

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1242 SAS No.: Q1242 SDG NO.: Q1242

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 01/14/2025 01/14/2025

Client Sample No.: CCAL02 Date Analyzed: 02/03/2025

Lab Sample No.: HSTDCCC750 Data File : PS029060.D Time Analyzed: 20:57

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.180	9.089	9.289	720.700	712.500	1.2
2,4-D	8.308	8.216	8.416	699.060	705.000	-0.8
2,4-DCAA	7.192	7.097	7.297	741.660	750.000	-1.1

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1242 SAS No.: Q1242 SDG NO.: Q1242

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/14/2025 01/14/2025

Client Sample No.: CCAL02 Date Analyzed: 02/03/2025

Lab Sample No.: HSTDCCC750 Data File : PS029060.D Time Analyzed: 20:57

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.803	9.713	9.913	710.800	712.500	-0.2
2,4-D	8.903	8.813	9.013	657.380	705.000	-6.8
2,4-DCAA	7.669	7.577	7.777	680.490	750.000	-9.3