

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS011425

Review By	Abdul	Review On	1/14/2025 4:08:46 PM
Supervise By	Ankita	Supervise On	1/15/2025 7:45:27 AM
SubDirectory	PS011425	HP Acquire Method	HP Processing Method ps011425 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24064,PP24065,PP24066,PP24067,PP24068
CCC Internal Standard/PEM	PP24066
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24069,PP24070

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS028899.D	14 Jan 2025 09:43		AR\AJ	Ok
2	I.BLK	I.BLK	PS028900.D	14 Jan 2025 10:07		AR\AJ	Ok
3	HSTDICC200	HSTDICC200	PS028901.D	14 Jan 2025 10:31		AR\AJ	Ok
4	HSTDICC500	HSTDICC500	PS028902.D	14 Jan 2025 10:55		AR\AJ	Ok
5	HSTDICC750	HSTDICC750	PS028903.D	14 Jan 2025 11:19		AR\AJ	Ok
6	HSTDICC1000	HSTDICC1000	PS028904.D	14 Jan 2025 11:43		AR\AJ	Ok
7	HSTDICC1500	HSTDICC1500	PS028905.D	14 Jan 2025 12:07		AR\AJ	Ok,M
8	HSTDICV750	ICVPS011425	PS028906.D	14 Jan 2025 12:31		AR\AJ	Ok,M
9	I.BLK	I.BLK	PS028907.D	14 Jan 2025 12:56		AR\AJ	Ok
10	HSTDCCC750	HSTDCCC750	PS028908.D	14 Jan 2025 13:20		AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS020325

Review By	Abdul	Review On	2/4/2025 8:13:08 AM
Supervise By	Ankita	Supervise On	2/4/2025 9:01:51 AM
SubDirectory	PS020325	HP Acquire Method	HP Processing Method ps114225 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24064,PP24065,PP24066,PP24067,PP24068
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24066 PP24069,PP24070

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS029045.D	03 Feb 2025 09:50		AR\AJ	Ok
2	I.BLK	I.BLK	PS029046.D	03 Feb 2025 10:14		AR\AJ	Ok
3	HSTDCCC750	HSTDCCC750	PS029047.D	03 Feb 2025 10:38		AR\AJ	Ok
4	PB166485BL	PB166485BL	PS029048.D	03 Feb 2025 16:09		AR\AJ	Ok
5	PB166485BS	PB166485BS	PS029049.D	03 Feb 2025 16:33		AR\AJ	Ok
6	PB166423TB	PB166423TB	PS029050.D	03 Feb 2025 16:57		AR\AJ	Ok
7	Q1241-04	JPP-3.5-013025	PS029051.D	03 Feb 2025 17:21		AR\AJ	Ok
8	Q1241-04MS	JPP-3.5-013025MS	PS029052.D	03 Feb 2025 17:45	Comp#14 recovery fail	AR\AJ	Ok,M
9	Q1241-04MSD	JPP-3.5-013025MSD	PS029053.D	03 Feb 2025 18:09	Comp#1,14 recovery fail , RPD fail	AR\AJ	Ok,M
10	Q1241-08	JPP-5.3-013025	PS029054.D	03 Feb 2025 18:33		AR\AJ	Ok
11	Q1241-12	JPP-5.2-013025	PS029055.D	03 Feb 2025 18:57		AR\AJ	Ok
12	Q1241-16	JPP-5.4-013025	PS029056.D	03 Feb 2025 19:21		AR\AJ	Ok
13	Q1241-20	JPP-5.1.4-013025	PS029057.D	03 Feb 2025 19:45		AR\AJ	Ok
14	Q1242-04	JPP-6.2-013025	PS029058.D	03 Feb 2025 20:09		AR\AJ	Ok
15	I.BLK	I.BLK	PS029059.D	03 Feb 2025 20:33		AR\AJ	Ok
16	HSTDCCC750	HSTDCCC750	PS029060.D	03 Feb 2025 20:57		AR\AJ	Ok

M : Manual Integration