

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1242

Client: RU2 Engineering, LLC

Analytical Method: 8151A

DataFile : PS029053.D

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID:	JPP-3,5-013025MSD											
Q1241-04MSD	2,4-D	50	0	53.3	ug/L	107		1		65	135	20
	2,4,5-TP(Silvex)	50	0	51.7	ug/L	103		1		62	139	20