

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1242 SAS No.: Q1242 SDG NO.: Q1242

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL04 Date Analyzed: 02/04/2025

Lab Sample No.: PSTDCCC050 Data File : PL094052.D Time Analyzed: 14:50

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.055	8.953	9.153	45.130	50.000	-9.7
Endrin	6.572	6.472	6.672	44.960	50.000	-10.1
gamma-BHC (Lindane)	4.327	4.227	4.427	50.110	50.000	0.2
Heptachlor	4.915	4.814	5.014	47.450	50.000	-5.1
Heptachlor epoxide	5.683	5.582	5.782	48.530	50.000	-2.9
Methoxychlor	7.499	7.398	7.598	43.350	50.000	-13.3
Tetrachloro-m-xylene	3.538	3.439	3.639	50.380	50.000	0.8

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1242 SAS No.: Q1242 SDG NO.: Q1242

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL04 Date Analyzed: 02/04/2025

Lab Sample No.: PSTDCCC050 Data File : PL094052.D Time Analyzed: 14:50

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.909	7.810	8.010	43.570	50.000	-12.9
Endrin	5.636	5.536	5.736	50.920	50.000	1.8
gamma-BHC (Lindane)	3.606	3.507	3.707	52.180	50.000	4.4
Heptachlor	3.944	3.845	4.045	48.050	50.000	-3.9
Heptachlor epoxide	4.726	4.627	4.827	51.700	50.000	3.4
Methoxychlor	6.609	6.509	6.709	42.000	50.000	-16.0
Tetrachloro-m-xylene	2.774	2.674	2.874	52.500	50.000	5.0