

Surrogate Summary

SDG No.: Q1242

Client: RU2 Engineering, LLC

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PL093725.D	PIBLK-PL093725.D	Decachlorobiphenyl	1	20	22.1	111		43	140
		Tetrachloro-m-xylene	1	20	20.8	104		77	126
		Decachlorobiphenyl	2	20	21.9	109		43	140
		Tetrachloro-m-xylene	2	20	20.5	103		77	126
I.BLK-PL093999.D	PIBLK-PL093999.D	Decachlorobiphenyl	1	20	28.1	140		43	140
		Tetrachloro-m-xylene	1	20	21.1	106		77	126
		Decachlorobiphenyl	2	20	21.9	109		43	140
		Tetrachloro-m-xylene	2	20	20.5	102		77	126
PB166484BL	PB166484BL	Decachlorobiphenyl	1	20	24.0	120		43	140
		Tetrachloro-m-xylene	1	20	23.4	117		77	126
		Decachlorobiphenyl	2	20	24.6	123		43	140
		Tetrachloro-m-xylene	2	20	22.3	112		77	126
PB166423TB	PB166423TB	Decachlorobiphenyl	1	20	24.6	123		43	140
		Tetrachloro-m-xylene	1	20	23.7	118		77	126
		Decachlorobiphenyl	2	20	25.9	130		43	140
		Tetrachloro-m-xylene	2	20	22.7	114		77	126
Q1241-04MS	JPP-3.5-013025MS	Decachlorobiphenyl	1	20	25.2	126		43	140
		Tetrachloro-m-xylene	1	20	22.9	114		77	126
		Decachlorobiphenyl	2	20	26.6	133		43	140
		Tetrachloro-m-xylene	2	20	22.3	112		77	126
Q1241-04MSD	JPP-3.5-013025MSD	Decachlorobiphenyl	1	20	25.1	126		43	140
		Tetrachloro-m-xylene	1	20	23.1	116		77	126
		Decachlorobiphenyl	2	20	26.7	134		43	140
		Tetrachloro-m-xylene	2	20	22.4	112		77	126
Q1242-04	JPP-6.2-013025	Decachlorobiphenyl	1	20	23.8	119		43	140
		Tetrachloro-m-xylene	1	20	23.9	119		77	126
		Decachlorobiphenyl	2	20	25.4	127		43	140
		Tetrachloro-m-xylene	2	20	23.3	116		77	126
I.BLK-PL094020.D	PIBLK-PL094020.D	Decachlorobiphenyl	1	20	22.8	114		43	140
		Tetrachloro-m-xylene	1	20	21.6	108		77	126
		Decachlorobiphenyl	2	20	24.3	121		43	140
		Tetrachloro-m-xylene	2	20	21.1	106		77	126
I.BLK-PL094036.D	PIBLK-PL094036.D	Decachlorobiphenyl	1	20	23.0	115		43	140
		Tetrachloro-m-xylene	1	20	22.4	112		77	126
		Decachlorobiphenyl	2	20	23.3	116		43	140
		Tetrachloro-m-xylene	2	20	21.9	110		77	126
PB166484BS	PB166484BS	Decachlorobiphenyl	1	20	20.7	103		43	140
		Tetrachloro-m-xylene	1	20	21.6	108		77	126
		Decachlorobiphenyl	2	20	19.1	95		43	140
		Tetrachloro-m-xylene	2	20	20.4	102		77	126
I.BLK-PL094051.D	PIBLK-PL094051.D	Decachlorobiphenyl	1	20	20.1	101		43	140

Surrogate Summary

SDG No.: Q1242

Client: RU2 Engineering, LLC

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PL094051.D	PIBLK-PL094051.D	Tetrachloro-m-xylene	1	20	21.9	109		77	126
		Decachlorobiphenyl	2	20	19.6	98		43	140
		Tetrachloro-m-xylene	2	20	21.4	107		77	126