

### Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1242

Client: RU2 Engineering, LLC

Analytical Method: 8081B

DataFile : PL094013.D

| Lab Sample ID:                            | Parameter           | Spike | Sample<br>Result | Result | Units | Rec<br>Qual | RPD<br>Qual | Low | Limits<br>High | RPD |
|-------------------------------------------|---------------------|-------|------------------|--------|-------|-------------|-------------|-----|----------------|-----|
| <b>Client Sample ID: JPP-3.5-013025MS</b> |                     |       |                  |        |       |             |             |     |                |     |
| Q1241-04MS                                | gamma-BHC (Lindane) | 5     | 0                | 5.00   | ug/L  | 100         |             | 60  | 152            |     |
|                                           | Heptachlor          | 5     | 0                | 5.30   | ug/L  | 106         |             | 56  | 147            |     |
|                                           | Heptachlor epoxide  | 5     | 0                | 5.10   | ug/L  | 102         |             | 77  | 143            |     |
|                                           | Endrin              | 5     | 0                | 5.60   | ug/L  | 112         |             | 76  | 144            |     |
|                                           | Methoxychlor        | 5     | 0                | 5.80   | ug/L  | 116         |             | 70  | 142            |     |