

Surrogate Summary

SW-846

SDG No.: Q1242

Client: RU2 Engineering, LLC

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB166418BL	PB166418BL	2-Fluorophenol	150	134	89		18	112
		Phenol-d6	150	129	86		15	107
		Nitrobenzene-d5	100	96.1	96		18	107
		2-Fluorobiphenyl	100	96.5	97		20	109
		2,4,6-Tribromophenol	150	147	98		10	116
		Terphenyl-d14	100	94.3	94		10	105
PB166418BS	PB166418BS	2-Fluorophenol	150	139	92		18	112
		Phenol-d6	150	133	89		15	107
		Nitrobenzene-d5	100	97.1	97		18	107
		2-Fluorobiphenyl	100	98.9	99		20	109
		2,4,6-Tribromophenol	150	160	106		10	116
		Terphenyl-d14	100	96.1	96		10	105
Q1239-10MS	357MS	2-Fluorophenol	150	80.0	53		18	112
		Phenol-d6	150	78.3	52		15	107
		Nitrobenzene-d5	100	54.3	54		18	107
		2-Fluorobiphenyl	100	53.2	53		20	109
		2,4,6-Tribromophenol	150	80.6	54		10	116
		Terphenyl-d14	100	46.0	46		10	105
Q1239-10MSD	357MSD	2-Fluorophenol	150	84.0	56		18	112
		Phenol-d6	150	82.0	55		15	107
		Nitrobenzene-d5	100	57.9	58		18	107
		2-Fluorobiphenyl	100	55.5	55		20	109
		2,4,6-Tribromophenol	150	84.1	56		10	116
		Terphenyl-d14	100	49.7	50		10	105
Q1242-03	JPP-6.2-013025	2-Fluorophenol	150	11.5	8	*	18	112
		Phenol-d6	150	61.9	41		15	107
		Nitrobenzene-d5	100	60.4	60		18	107
		2-Fluorobiphenyl	100	65.5	65		20	109
		2,4,6-Tribromophenol	150	5.80	4	*	10	116
		Terphenyl-d14	100	66.5	67		10	105