

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1242

Client: RU2 Engineering, LLC

Analytical Method: 8270E

DataFile: BF141441.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB166418BS	Benzaldehyde	1700	1700	ug/Kg	100				10	133	
	Phenol	1700	1400	ug/Kg	82				62	112	
	bis(2-Chloroethyl)ether	1700	1400	ug/Kg	82				60	101	
	2-Chlorophenol	1700	1500	ug/Kg	88				65	112	
	2-Methylphenol	1700	1400	ug/Kg	82				61	108	
	2,2-oxybis(1-Chloropropane)	1700	1300	ug/Kg	76				51	100	
	Acetophenone	1700	1700	ug/Kg	100		*		66	98	
	3+4-Methylphenols	1700	1500	ug/Kg	88				58	111	
	N-Nitroso-di-n-propylamine	1700	1400	ug/Kg	82				63	95	
	Hexachloroethane	1700	1500	ug/Kg	88				72	108	
	Nitrobenzene	1700	1400	ug/Kg	82				57	101	
	Isophorone	1700	1500	ug/Kg	88				59	99	
	2-Nitrophenol	1700	1500	ug/Kg	88				61	111	
	2,4-Dimethylphenol	1700	1800	ug/Kg	106				46	141	
	bis(2-Chloroethoxy)methane	1700	1300	ug/Kg	76				66	97	
	2,4-Dichlorophenol	1700	1500	ug/Kg	88				62	107	
	Naphthalene	1700	1400	ug/Kg	82				62	100	
	4-Chloroaniline	1700	580	ug/Kg	34				16	100	
	Hexachlorobutadiene	1700	1500	ug/Kg	88				53	98	
	Caprolactam	1700	1600	ug/Kg	94				67	110	
	4-Chloro-3-methylphenol	1700	1500	ug/Kg	88				58	112	
	2-Methylnaphthalene	1700	1400	ug/Kg	82				60	104	
	Hexachlorocyclopentadiene	3300	6300	ug/Kg	191		*		45	165	
	2,4,6-Trichlorophenol	1700	1500	ug/Kg	88				59	102	
	2,4,5-Trichlorophenol	1700	1400	ug/Kg	82				61	98	
	1,1-Biphenyl	1700	1600	ug/Kg	94				57	103	
	2-Chloronaphthalene	1700	1400	ug/Kg	82				58	99	
	2-Nitroaniline	1700	1500	ug/Kg	88				66	101	
	Dimethylphthalate	1700	1500	ug/Kg	88				61	99	
	Acenaphthylene	1700	1600	ug/Kg	94				63	101	
	2,6-Dinitrotoluene	1700	1400	ug/Kg	82				61	104	
	3-Nitroaniline	1700	750	ug/Kg	44				28	100	
	Acenaphthene	1700	1700	ug/Kg	100				57	104	
	2,4-Dinitrophenol	3300	3700	ug/Kg	112				37	128	
	4-Nitrophenol	3300	3200	ug/Kg	97				48	119	
	Dibenzofuran	1700	1500	ug/Kg	88				63	99	
	2,4-Dinitrotoluene	1700	1500	ug/Kg	88				60	106	
	Diethylphthalate	1700	1500	ug/Kg	88				60	101	
	4-Chlorophenyl-phenylether	1700	1500	ug/Kg	88				58	98	
	Fluorene	1700	1500	ug/Kg	88				61	101	
	4-Nitroaniline	1700	1500	ug/Kg	88				64	103	
	4,6-Dinitro-2-methylphenol	1700	1700	ug/Kg	100				76	113	
	N-Nitrosodiphenylamine	1700	1500	ug/Kg	88				71	99	
	4-Bromophenyl-phenylether	1700	1500	ug/Kg	88				66	102	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1242

Client: RU2 Engineering, LLC

Analytical Method: 8270E

DataFile: BF141441.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD		Limits		RPD
								Qual	Low	High		
PB166418BS	Hexachlorobenzene	1700	1500	ug/Kg	88				64	98		
	Atrazine	1700	1700	ug/Kg	100				47	152		
	Pentachlorophenol	3300	3100	ug/Kg	94				67	105		
	Phenanthrene	1700	1500	ug/Kg	88				59	103		
	Anthracene	1700	1600	ug/Kg	94				61	105		
	Carbazole	1700	1600	ug/Kg	94				61	99		
	Di-n-butylphthalate	1700	1500	ug/Kg	88				58	104		
	Fluoranthene	1700	1600	ug/Kg	94				57	107		
	Pyrene	1700	1400	ug/Kg	82				59	103		
	Butylbenzylphthalate	1700	1500	ug/Kg	88				55	103		
	3,3-Dichlorobenzidine	1700	710	ug/Kg	42				42	91		
	Benzo(a)anthracene	1700	1500	ug/Kg	88				60	102		
	Chrysene	1700	1400	ug/Kg	82				59	101		
	bis(2-Ethylhexyl)phthalate	1700	1500	ug/Kg	88				54	135		
	Di-n-octyl phthalate	1700	1400	ug/Kg	82				52	137		
	Benzo(b)fluoranthene	1700	1400	ug/Kg	82				62	109		
	Benzo(k)fluoranthene	1700	1600	ug/Kg	94				62	109		
	Benzo(a)pyrene	1700	1600	ug/Kg	94				63	103		
	Indeno(1,2,3-cd)pyrene	1700	1500	ug/Kg	88				63	101		
	Dibenz(a,h)anthracene	1700	1500	ug/Kg	88				61	112		
	Benzo(g,h,i)perylene	1700	1400	ug/Kg	82				70	108		
	1,2,4,5-Tetrachlorobenzene	1700	1600	ug/Kg	94				53	101		
	1,4-Dioxane	1700	1300	ug/Kg	76				50	96		
	2,3,4,6-Tetrachlorophenol	1700	1600	ug/Kg	94				59	108		