

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1242 SAS No.: Q1242 SDG NO.: Q1242

EPA Sample No.: SSTDCCC040 Date Analyzed: 02/04/2025

Lab File ID: BF141398.D Time Analyzed: 10:42

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	87552	6.793	331835	8.08	180473	9.83
UPPER LIMIT	175104	7.293	663670	8.575	360946	10.328
LOWER LIMIT	43776	6.293	165918	7.575	90236.5	9.328
EPA SAMPLE NO.						
01 PB166483BL	74401	6.79	304159	8.07	164089	9.83
02 PB166483BS	76563	6.79	300075	8.08	160616	9.83
03 PB166423TB	79603	6.79	321231	8.07	179300	9.83

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.