

### VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: RUTW01

Lab Code: CHEM Case No.: Q1242 SAS No.: Q1242 SDG No.: Q1242

Instrument ID: MSVOA\_Y Calibration Date(s): 02/03/2025 02/03/2025

Heated Purge: (Y/N) Y Calibration Time(s): 10:35 12:44

GC Column: RXI-624 ID: 0.25 (mm)

| LAB FILE ID:                   |  | RRF005 = VY021020.D |        | RRF010 = VY021021.D |        | RRF020 = VY021022.D |        | RRF050 = VY021023.D |       | RRF100 = VY021024.D |  | RRF150 = VY021025.D |  |
|--------------------------------|--|---------------------|--------|---------------------|--------|---------------------|--------|---------------------|-------|---------------------|--|---------------------|--|
| COMPOUND                       |  | RRF005              | RRF010 | RRF020              | RRF050 | RRF100              | RRF150 | RRF                 | % RSD |                     |  |                     |  |
| Dichlorodifluoromethane        |  | 0.503               | 0.438  | 0.413               | 0.445  | 0.419               | 0.457  | 0.446               | 7.2   |                     |  |                     |  |
| Chloromethane                  |  | 0.524               | 0.445  | 0.408               | 0.421  | 0.404               | 0.437  | 0.440               | 10.1  |                     |  |                     |  |
| Vinyl Chloride                 |  | 0.535               | 0.441  | 0.404               | 0.424  | 0.408               | 0.443  | 0.443               | 10.9  |                     |  |                     |  |
| Bromomethane                   |  | 0.356               | 0.275  | 0.249               | 0.258  | 0.246               | 0.257  | 0.274               | 15.2  |                     |  |                     |  |
| Chloroethane                   |  | 0.316               | 0.270  | 0.240               | 0.255  | 0.239               | 0.261  | 0.263               | 10.7  |                     |  |                     |  |
| Trichlorofluoromethane         |  | 0.924               | 0.814  | 0.755               | 0.789  | 0.743               | 0.813  | 0.806               | 8     |                     |  |                     |  |
| 1,1,2-Trichlorotrifluoroethane |  | 0.587               | 0.507  | 0.472               | 0.494  | 0.466               | 0.510  | 0.506               | 8.6   |                     |  |                     |  |
| Tert butyl alcohol             |  | 0.042               | 0.039  | 0.035               | 0.032  | 0.028               | 0.032  | 0.035               | 14.8  |                     |  |                     |  |
| 1,1-Dichloroethene             |  | 0.529               | 0.471  | 0.441               | 0.469  | 0.452               | 0.498  | 0.477               | 6.8   |                     |  |                     |  |
| Acetone                        |  | 0.087               | 0.083  | 0.078               | 0.077  | 0.070               | 0.083  | 0.080               | 7.5   |                     |  |                     |  |
| Carbon Disulfide               |  | 1.673               | 1.456  | 1.389               | 1.475  | 1.421               | 1.556  | 1.495               | 7     |                     |  |                     |  |
| Methyl tert-butyl Ether        |  | 1.271               | 1.185  | 1.168               | 1.216  | 1.124               | 1.265  | 1.205               | 4.8   |                     |  |                     |  |
| Methyl Acetate                 |  | 0.300               | 0.282  | 0.268               | 0.287  | 0.253               | 0.291  | 0.280               | 6.1   |                     |  |                     |  |
| Methylene Chloride             |  | 0.570               | 0.506  | 0.473               | 0.485  | 0.452               | 0.494  | 0.497               | 8.2   |                     |  |                     |  |
| trans-1,2-Dichloroethene       |  | 0.578               | 0.518  | 0.502               | 0.522  | 0.490               | 0.538  | 0.525               | 5.8   |                     |  |                     |  |
| 1,1-Dichloroethane             |  | 1.104               | 0.948  | 0.923               | 0.947  | 0.911               | 0.989  | 0.970               | 7.3   |                     |  |                     |  |
| Cyclohexane                    |  | 1.093               | 0.885  | 0.807               | 0.827  | 0.779               | 0.852  | 0.874               | 13    |                     |  |                     |  |
| 2-Butanone                     |  | 0.134               | 0.133  | 0.126               | 0.135  | 0.121               | 0.139  | 0.131               | 5.2   |                     |  |                     |  |
| Carbon Tetrachloride           |  | 0.604               | 0.551  | 0.522               | 0.547  | 0.522               | 0.581  | 0.555               | 5.9   |                     |  |                     |  |
| cis-1,2-Dichloroethene         |  | 0.662               | 0.584  | 0.577               | 0.599  | 0.575               | 0.626  | 0.604               | 5.7   |                     |  |                     |  |
| Bromochloromethane             |  | 0.373               | 0.368  | 0.381               | 0.413  | 0.374               | 0.394  | 0.384               | 4.4   |                     |  |                     |  |
| Chloroform                     |  | 1.107               | 1.016  | 0.947               | 0.982  | 0.921               | 1.006  | 0.997               | 6.5   |                     |  |                     |  |
| 1,1,1-Trichloroethane          |  | 1.067               | 0.901  | 0.867               | 0.912  | 0.863               | 0.949  | 0.927               | 8.2   |                     |  |                     |  |
| Methylcyclohexane              |  | 0.609               | 0.550  | 0.536               | 0.588  | 0.570               | 0.629  | 0.581               | 6.1   |                     |  |                     |  |
| Benzene                        |  | 1.512               | 1.382  | 1.327               | 1.378  | 1.323               | 1.436  | 1.393               | 5.1   |                     |  |                     |  |
| 1,2-Dichloroethane             |  | 0.435               | 0.390  | 0.367               | 0.379  | 0.352               | 0.395  | 0.386               | 7.4   |                     |  |                     |  |
| Trichloroethene                |  | 0.401               | 0.355  | 0.341               | 0.349  | 0.338               | 0.375  | 0.360               | 6.7   |                     |  |                     |  |
| 1,2-Dichloropropane            |  | 0.357               | 0.329  | 0.319               | 0.327  | 0.311               | 0.341  | 0.331               | 5     |                     |  |                     |  |
| Bromodichloromethane           |  | 0.528               | 0.500  | 0.469               | 0.482  | 0.459               | 0.506  | 0.491               | 5.2   |                     |  |                     |  |
| 4-Methyl-2-Pentanone           |  | 0.202               | 0.211  | 0.202               | 0.216  | 0.194               | 0.224  | 0.208               | 5.2   |                     |  |                     |  |

\* Compounds with required minimum RRF and maximum %RSD values.  
All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.

### VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: RUTW01  
 Lab Code: CHEM Case No.: Q1242 SAS No.: Q1242 SDG No.: Q1242  
 Instrument ID: MSVOA\_Y Calibration Date(s): 02/03/2025 02/03/2025  
 Heated Purge: (Y/N) Y Calibration Time(s): 10:35 12:44  
 GC Column: RXI-624 ID: 0.25 (mm)

| LAB FILE ID:                |        |        |                     |        |        |                     |       |       |
|-----------------------------|--------|--------|---------------------|--------|--------|---------------------|-------|-------|
| RRF005 = VY021020.D         |        |        | RRF010 = VY021021.D |        |        | RRF020 = VY021022.D |       |       |
| RRF050 = VY021023.D         |        |        | RRF100 = VY021024.D |        |        | RRF150 = VY021025.D |       |       |
| COMPOUND                    | RRF005 | RRF010 | RRF020              | RRF050 | RRF100 | RRF150              | RRF   | % RSD |
| Toluene                     | 0.913  | 0.856  | 0.840               | 0.876  | 0.844  | 0.918               | 0.875 | 3.9   |
| t-1,3-Dichloropropene       | 0.441  | 0.426  | 0.419               | 0.449  | 0.427  | 0.483               | 0.441 | 5.3   |
| cis-1,3-Dichloropropene     | 0.528  | 0.516  | 0.502               | 0.525  | 0.501  | 0.560               | 0.522 | 4.2   |
| 1,1,2-Trichloroethane       | 0.261  | 0.235  | 0.231               | 0.235  | 0.217  | 0.242               | 0.237 | 6.1   |
| 2-Hexanone                  | 0.124  | 0.131  | 0.128               | 0.144  | 0.129  | 0.149               | 0.134 | 7.4   |
| Dibromochloromethane        | 0.359  | 0.330  | 0.324               | 0.329  | 0.310  | 0.346               | 0.333 | 5.2   |
| 1,2-Dibromoethane           | 0.232  | 0.223  | 0.215               | 0.225  | 0.203  | 0.233               | 0.222 | 5     |
| Tetrachloroethene           | 0.409  | 0.367  | 0.352               | 0.361  | 0.351  | 0.387               | 0.371 | 6.2   |
| Chlorobenzene               | 1.237  | 1.115  | 1.054               | 1.080  | 1.048  | 1.146               | 1.113 | 6.4   |
| Ethyl Benzene               | 2.101  | 1.899  | 1.865               | 1.943  | 1.914  | 2.093               | 1.969 | 5.2   |
| m/p-Xylenes                 | 0.772  | 0.723  | 0.709               | 0.733  | 0.711  | 0.773               | 0.737 | 3.9   |
| o-Xylene                    | 0.723  | 0.658  | 0.660               | 0.681  | 0.670  | 0.729               | 0.687 | 4.6   |
| Styrene                     | 1.154  | 1.118  | 1.105               | 1.156  | 1.128  | 1.208               | 1.145 | 3.2   |
| Bromoform                   | 0.231  | 0.222  | 0.217               | 0.216  | 0.205  | 0.229               | 0.220 | 4.4   |
| Isopropylbenzene            | 4.122  | 3.695  | 3.673               | 3.809  | 3.871  | 4.263               | 3.906 | 6.1   |
| 1,1,2,2-Tetrachloroethane   | 0.707  | 0.646  | 0.635               | 0.632  | 0.597  | 0.680               | 0.650 | 6     |
| 1,3-Dichlorobenzene         | 1.937  | 1.686  | 1.660               | 1.683  | 1.652  | 1.824               | 1.740 | 6.6   |
| 1,4-Dichlorobenzene         | 1.906  | 1.702  | 1.632               | 1.665  | 1.599  | 1.778               | 1.714 | 6.6   |
| 1,2-Dichlorobenzene         | 1.668  | 1.524  | 1.440               | 1.460  | 1.405  | 1.567               | 1.510 | 6.4   |
| 1,2-Dibromo-3-Chloropropane | 0.099  | 0.099  | 0.089               | 0.097  | 0.089  | 0.109               | 0.097 | 7.6   |
| 1,2,4-Trichlorobenzene      | 0.796  | 0.754  | 0.743               | 0.875  | 0.873  | 1.004               | 0.841 | 11.6  |
| 1,2,3-Trichlorobenzene      | 0.641  | 0.640  | 0.626               | 0.733  | 0.726  | 0.832               | 0.700 | 11.4  |
| 1,2-Dichloroethane-d4       | 0.566  | 0.557  | 0.478               | 0.503  | 0.454  | 0.516               | 0.512 | 8.5   |
| Dibromofluoromethane        | 0.351  | 0.326  | 0.301               | 0.313  | 0.297  | 0.333               | 0.320 | 6.4   |
| Toluene-d8                  | 1.295  | 1.237  | 1.139               | 1.202  | 1.158  | 1.288               | 1.220 | 5.3   |
| 4-Bromofluorobenzene        | 0.417  | 0.408  | 0.371               | 0.403  | 0.375  | 0.418               | 0.399 | 5.2   |

\* Compounds with required minimum RRF and maximum %RSD values.  
 All other compounds must meet a minimum RRF of 0.010.  
 RRF of 1,4-Dioxane = Value should be divide by 1000.