EPA	SAMPLE	NO.

A6335A

FORM 5B - IN POST-DIGESTION/DISTILLATION SPIKE SAMPLE RECOVERY

Lab Name: Alliance Technical Group, LLC			Contract:	68HERH20D0011			
Lab Code:	ACE	Case No.:	51952	MA No. :		SDG No.:	A6311
Matrix : Water			Analytical	Method:	ICP-MS		
% Solids:							

Concentration Units (µg/L, mg/L or mg/kg dry weight): _____ug/L ____

Analyte	Control Limit %R	Spiked Sample Result (SSR)	Q	Sample Result (SR)	Q	Spike Added (SA)	%R	Q
Selenium	75 - 125	13		5.0	U	10.0	127	