



A Waters Company



Instructions for Catalog # 500QR Water Pollution Trace Metals

Revision 120823

Description:

- This standard is packaged in a 30 mL screw-top vial containing approximately 28 mL of standard concentrate.
- This concentrate is preserved with approximately 2% (v/v) nitric acid and 0.1% (w/v) tartaric acid.
- The concentrate can be stored at room temperature.
- The diluted standard will contain all the analytes listed in the ranges specified on the data reporting form.

Before you begin:

- The sample resulting from the dilution described below will have a nitric acid concentration of approximately 0.02% before any acid is added. You may add a volume of acid different from the 2 to 5 mL of HNO_3 suggested in order to matrix match your calibration standards or meet any other method criteria.
- If analyzing this standard using colorimetric techniques, it may be necessary to pH adjust the sample prior to analysis. If using colorimetric techniques, it is acceptable to omit the addition of the 2-5 mL nitric acid suggested.
- While it is technically not necessary to digest this standard prior to analysis, digestion should be performed if this is your normal procedure.
- This standard should be analyzed as soon as possible after the concentrate is diluted.

Instructions:

1. Add 100-200 mL of deionized water and approximately 2 to 5 mL of nitric acid to a clean 500 mL class A volumetric flask.
2. Shake the Trace Metals vial prior to opening.
3. Using a clean, dry, class A pipet, volumetrically pipet 10.0 mL of the concentrate into the 500 mL volumetric flask.
4. Dilute the flask to final volume (500 mL) with deionized water.
5. Cap the flask and mix well.
6. Immediately analyze the diluted sample by your normal procedures.
7. Report your results as $\mu\text{g/L}$ for the diluted sample.

Safety:

ERA products may be hazardous and are intended for use by professional laboratory personnel trained in the competent handling of such materials. Responsibility for the safe use of these products rests entirely with the buyer and/or user. Safety Data Sheets (SDS) for all ERA products are available through our website.
www.eraqc.com



A Waters Company



Instructions for Catalog #666QR

WatR™ Supply Mercury

Revision 030512

Description:

- This standard is packaged in a 15 mL screw-top vial containing approximately 14 mL of standard concentrate.
- This concentrate is preserved with approximately 1% (v/v) nitric acid and 0.04% (w/v) potassium dichromate.
- The concentrate can be stored at room temperature.
- The diluted standard will contain Mercury in the range specified on the data reporting form.

Before you begin:

- This standard has been prepared as a concentrate and must be diluted prior to analysis.
- The USEPA *Criteria Document* requires that Mercury be present as a mixture of organic and inorganic forms and must, therefore, be analyzed as Total Mercury.
- This standard should be analyzed as soon as possible after the concentrate is diluted.

Instructions:

1. Add 100-200 mL of deionized water and approximately 2 to 5 mL of nitric acid to a clean 500 mL class A volumetric flask.
2. Shake the Mercury vial prior to opening.
3. Using a clean, dry, class A pipet, volumetrically pipet 5.0 mL of the concentrate into the 500 mL volumetric flask.
4. Dilute the flask to final volume with deionized water.
5. Cap the flask and mix well.
6. Immediately analyze the diluted sample by your normal procedures.
7. Report your results as $\mu\text{g/L}$ for the diluted sample.

Safety:

ERA products may be hazardous and are intended for use by professional laboratory personnel trained in the competent handling of such materials. Responsibility for the safe use of these products rests entirely with the buyer and/or user. Material Safety Data Sheets (MSDS) for all ERA products are available by calling 1-800-372-0122.