

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51952 MA No. : SDG No.: A6313
 Initial Calibration Verification Source : MP84212
 Continuing Calibration Verification Source : MP84215
 Run Batch: LB134523 Analytical Method: ICP-AES
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV007				ID: CCV033				ID: CCV034		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum	2500	2500	99	0	400000	370000	93	1	400000	101	1
Calcium	10000	10000	100	1	400000	380000	94	0	390000	99	1
Iron	10000	9400	94	1	400000	380000	94	0	400000	101	0
Magnesium	6000	6300	104	0	400000	370000	93	0	390000	98	0
Potassium	9900	9400	95	1	150000	140000	96	0	150000	103	1
Sodium	10000	10000	104	2	400000	380000	95	0	360000	90	2

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51952 MA No. : SDG No.: A6313Initial Calibration Verification Source : MP84212Continuing Calibration Verification Source : MP84215Run Batch: LB134523 Analytical Method: ICP-AESConcentration Units: $\mu\text{g/L}$

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV035				ID:		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum					400000	420000	105	0			
Calcium					400000	420000	104	1			
Iron					400000	420000	106	1			
Magnesium					400000	410000	103	1			
Potassium					150000	160000	107	1			
Sodium					400000	400000	100	1			