

FORM 4 - IN
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51952 MA No. : SDG No.: A6313
Analytical Method: ICP-AES ICSA Source: EPA-1211
Instrument ID: P4 ICSB Source: EPA-0710
Run Batch: LB134523
Concentration Units: ug/L

Analyte	True		Found			
	ICSA	ICSAB	ICSA	%R	ICSAB	%R
Aluminum	255000	247000	220000	85	220000	90
Calcium	245000	235000	230000	93	230000	99
Iron	101000	99300	95000	94	97000	98
Magnesium	255000	248000	220000	87	230000	92
Potassium	0	0	110		61	
Sodium	0	0	-390		-550	