

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51952 MA No.: _____ SDG No.: A6313
Instrument ID: CV1 Start Date: 02/03/2025
Analytical Method: CVAA Run Batch: LB134528
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Mercury	0	0.062	0	0.2	0.19	-6	2.5	2.4	-5

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51952 MA No.: SDG No.: A6313
Instrument ID: CV1 Start Date: 02/03/2025
Analytical Method: CVAA Run Batch: LB134528
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Mercury	5	5.1	1	7.5	7.6	1	10	10	0