

LB134528


SFAM01.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

46417



A= 0.0000e+000
B= 2.1452e-004 slope
C= 5.0198e-004 y=mx+c
Rho= 0.9998348
Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	G/U)
0.00	0.000	0.062	0.062	285	0.000	285					7
0.05	0.050										-6
0.20	0.200	0.188	-0.012	874	0.0 %	874					-5
2.50	2.500	2.379	-0.121	11085	0.0 %	11085					1
5.00	5.000	5.062	0.062	23594	0.0 %	23594					1
7.50	7.500	7.552	0.052	35200	0.0 %	35200					1
10.0	10.000	9.958	-0.042	46417	0.0 %	46417					0

LB134528

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	285	-		0 SFAM01.1	PPB	2/3/2025 13:00	Std
	0.2 S01	874	-		0.2 SFAM01.1	PPB	2/3/2025 13:02	Std
	2.5 S02	11085	-		2.5 SFAM01.1	PPB	2/3/2025 13:05	Std
	5 S03	23594	-		5 SFAM01.1	PPB	2/3/2025 13:07	Std
	7.5 S04	35200	-		7.5 SFAM01.1	PPB	2/3/2025 13:09	Std
	10 S05	46417	-		10 SFAM01.1	PPB	2/3/2025 13:14	Std
ICV080	ICV080	18086	3.8804 -		SFAM01.1	PPB	2/3/2025 13:20	SMPL
ICB080	ICB080	-35	-0.007 -		SFAM01.1	PPB	2/3/2025 13:22	SMPL
CCV007	CCV007	23664	5.077 -		SFAM01.1	PPB	2/3/2025 13:24	SMPL
CCB007	CCB007	27	0.0063 -		SFAM01.1	PPB	2/3/2025 13:26	SMPL
PB166499BL	PBS499	217	0.0471 -		SFAM01.1	PPB	2/3/2025 13:29	SMPL
Q1163-01	ME2946	5299	1.1373 -		SFAM01.1	PPB	2/3/2025 13:31	SMPL
Q1163-02	ME2947	1282	0.2755 -		SFAM01.1	PPB	2/3/2025 13:33	SMPL
Q1163-03	ME2947D	1538	0.3304 -		SFAM01.1	PPB	2/3/2025 13:35	SMPL
Q1163-04	ME2947S	12973	2.7835 -		SFAM01.1	PPB	2/3/2025 13:42	SMPL
Q1246-01	A6313	208658	44.7628 -		SFAM01.1	PPB	2/3/2025 13:45	SMPL
Q1246-02	A6314	371900	HIGH -		SFAM01.1	PPB	2/3/2025 13:47	
Q1246-03	A6325	2566	0.551 -		SFAM01.1	PPB	2/3/2025 13:49	SMPL
Q1246-04	A6326	1320	0.2837 -		SFAM01.1	PPB	2/3/2025 13:51	SMPL
Q1246-05	A6331	4596	0.9865 -		SFAM01.1	PPB	2/3/2025 13:54	SMPL
Q1246-06	A6332	1400	0.3008 -		SFAM01.1	PPB	2/3/2025 13:56	SMPL
Q1246-07	A6333	5665	1.2158 -		SFAM01.1	PPB	2/3/2025 13:58	SMPL
Q1246-08	A6334	4683	1.0051 -		SFAM01.1	PPB	2/3/2025 14:01	SMPL
Q1246-09	A6342	807	0.1736 -		SFAM01.1	PPB	2/3/2025 14:03	SMPL
Q1246-10	A6342D	915	0.1968 -		SFAM01.1	PPB	2/3/2025 14:05	SMPL
Q1246-11	A6342S	12091	2.5943 -		SFAM01.1	PPB	2/3/2025 14:07	SMPL
Q1246-12	A6343	1171	0.2517 -		SFAM01.1	PPB	2/3/2025 14:10	SMPL
Q1246-13	A6344	850	0.1828 -		SFAM01.1	PPB	2/3/2025 14:12	SMPL
Q1246-14	A6345	685	0.1475 -		SFAM01.1	PPB	2/3/2025 14:14	SMPL
Q1246-15	A6346	540	0.1163 -		SFAM01.1	PPB	2/3/2025 14:16	SMPL
CCV008	CCV008	24221	5.1965 -		SFAM01.1	PPB	2/3/2025 14:19	SMPL
CCB008	CCB008	-36	-0.0072 -		SFAM01.1	PPB	2/3/2025 14:21	SMPL
Q1246-16	A6347	933	0.2007 -		SFAM01.1	PPB	2/3/2025 14:23	SMPL
Q1246-01DLX10	A6313	23939	5.136 -		SFAM01.1	PPB	2/3/2025 14:39	SMPL
Q1246-02DLX25	A6314	23691	5.0828 -		SFAM01.1	PPB	2/3/2025 14:53	SMPL
CCV009	CCV009	23484	5.0384 -		SFAM01.1	PPB	2/3/2025 14:57	SMPL
CCB009	CCB009	-91	-0.019 -		SFAM01.1	PPB	2/3/2025 15:00	SMPL