

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB134522

Review By	Review On
Supervise By	Supervise On

STD. NAME	STD REF.#
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard	

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	TUNE	TUNE001	TUNE	02/02/25 13:51		Jaswal	OK
2	S00	S0	CAL1	02/02/25 14:25		Jaswal	OK
3	S02	S02	CAL3	02/02/25 14:28		Jaswal	OK
4	S03	S03	CAL4	02/02/25 14:32		Jaswal	OK
5	S04	S04	CAL5	02/02/25 14:35		Jaswal	OK
6	S05	S05	CAL6	02/02/25 14:38		Jaswal	OK
7	S06	S06	CAL7	02/02/25 14:40		Jaswal	OK
8	S07	S07	CAL8	02/02/25 14:43		Jaswal	OK
9	S08	S08	CAL9	02/02/25 14:46		Jaswal	OK
10	ICV001	ICV001	ICV	02/02/25 14:51		Jaswal	OK
11	ICB001	ICB001	ICB	02/02/25 14:55		Jaswal	OK
12	ICSA001	ICSA001	ICSA	02/02/25 15:01	Inst. paused for review	Jaswal	OK
13	ICSAB001	ICSAB001	ICSAB	02/02/25 15:04		Jaswal	OK
14	CCV001	CCV001	CCV	02/02/25 15:07		Jaswal	OK
15	CCB001	CCB001	CCB	02/02/25 15:14	Inst. paused for review	Jaswal	OK
16	PB166384BL	PBS384	MB	02/02/25 15:17		Jaswal	OK
17	PB166384BS	LCS384	LCS	02/02/25 15:20		Jaswal	OK
18	PB166471BL	PBS471	MB	02/02/25 15:24		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB134522

Review By	Review On
Supervise By	Supervise On

STD. NAME	STD REF.#
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard	

19	PB166471BS	LCS471	LCS	02/02/25 15:27		Jaswal	OK
20	Q1223-05	A6305	SAM	02/02/25 15:30	Cr High	Jaswal	Dilution
21	Q1223-06	A6305D	DUP	02/02/25 15:33	Cr High	Jaswal	Dilution
22	Q1223-05L	A6305L	SD	02/02/25 15:36	Cr High	Jaswal	Dilution
23	Q1223-07	A6305S	MS	02/02/25 15:39	Cr High, MS Fail For Sb(Below RL),Se(Below RL),Zn	Jaswal	Dilution
24	Q1246-09	A6342	SAM	02/02/25 15:42		Jaswal	OK
25	Q1246-10	A6342D	DUP	02/02/25 15:46		Jaswal	OK
26	Q1246-09L	A6342L	SD	02/02/25 15:49		Jaswal	OK
27	Q1246-11	A6342S	MS	02/02/25 15:52	MS Fail For Se(Below RL)	Jaswal	OK
28	Q1223-08	A6306	SAM	02/02/25 15:55	Cr High	Jaswal	Dilution
29	Q1223-09	A6307	SAM	02/02/25 15:58		Jaswal	OK
30	Q1223-10	A6308	SAM	02/02/25 16:02	Cr High	Jaswal	Dilution
31	Q1223-11	A6309	SAM	02/02/25 16:05		Jaswal	OK
32	Q1223-12	A6320	SAM	02/02/25 16:08		Jaswal	OK
33	Q1223-13	A6321	SAM	02/02/25 16:11	Cr High	Jaswal	Dilution
34	Q1223-14	A6322	SAM	02/02/25 16:14	Cr High	Jaswal	Dilution
35	Q1223-15	A6323	SAM	02/02/25 16:17	Cr High	Jaswal	Dilution
36	Q1223-16	A6324	SAM	02/02/25 16:20		Jaswal	OK
37	CCV002	CCV002	CCV	02/02/25 16:24		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB134522

Review By	Review On
Supervise By	Supervise On

STD. NAME	STD REF.#
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard	

38	CCB002	CCB002	CCB	02/02/25 16:29		Jaswal	OK
39	Q1246-01	A6313	SAM	02/02/25 16:32		Jaswal	OK
40	Q1246-02	A6314	SAM	02/02/25 16:35		Jaswal	OK
41	Q1246-03	A6325	SAM	02/02/25 16:38		Jaswal	OK
42	Q1246-04	A6326	SAM	02/02/25 16:42		Jaswal	OK
43	Q1246-05	A6331	SAM	02/02/25 16:45		Jaswal	OK
44	Q1246-06	A6332	SAM	02/02/25 16:48		Jaswal	OK
45	Q1246-07	A6333	SAM	02/02/25 16:51		Jaswal	OK
46	Q1246-08	A6334	SAM	02/02/25 16:54		Jaswal	OK
47	Q1223-01	A6301	SAM	02/02/25 16:57		Jaswal	OK
48	Q1223-02	A6302	SAM	02/02/25 17:01	Cr,Mn High	Jaswal	Dilution
49	Q1223-03	A6303	SAM	02/02/25 17:04		Jaswal	OK
50	Q1223-04	A6304	SAM	02/02/25 17:07		Jaswal	OK
51	Q1246-12	A6343	SAM	02/02/25 17:10		Jaswal	OK
52	Q1246-13	A6344	SAM	02/02/25 17:13		Jaswal	OK
53	Q1246-14	A6345	SAM	02/02/25 17:16		Jaswal	OK
54	Q1246-15	A6346	SAM	02/02/25 17:20		Jaswal	OK
55	Q1246-16	A6347	SAM	02/02/25 17:23		Jaswal	OK
56	Q1223-05DL	A6305	SAM	02/02/25 17:26	2X For Cr	Jaswal	Confirms
57	Q1223-06DL	A6305D	DUP	02/02/25 17:29	2X For Cr	Jaswal	Confirms

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB134522

Review By	Review On
Supervise By	Supervise On

STD. NAME	STD REF.#
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard	

58	Q1223-05LDL	A6305L	SD	02/02/25 17:32	10X For Cr	Jaswal	Confirms
59	Q1223-07DL	A6305S	MS	02/02/25 17:35	2X For Cr	Jaswal	Confirms
60	Q1223-08DL	A6306	SAM	02/02/25 17:38	2X For Cr	Jaswal	Confirms
61	Q1223-10DL	A6308	SAM	02/02/25 17:42	25X For Cr	Jaswal	Confirms
62	Q1223-13DL	A6321	SAM	02/02/25 17:45	5X For Cr	Jaswal	Confirms
63	Q1223-14DL	A6322	SAM	02/02/25 17:48	5X For Cr	Jaswal	Confirms
64	Q1223-15DL	A6323	SAM	02/02/25 17:51	5X For Cr	Jaswal	Confirms
65	CCV003	CCV003	CCV	02/02/25 17:54		Jaswal	OK
66	CCB003	CCB003	CCB	02/02/25 17:57		Jaswal	OK
67	Q1223-02DL	A6302	SAM	02/02/25 18:00	2X For Cr,Mn Still Mn High	Jaswal	Dilution
68	Q1223-02DL2	A6302	SAM	02/02/25 18:04	5X For Mn	Jaswal	Confirms
69	Q1223-05A	A6305A	PS	02/02/25 18:07	PS For Sb,Se,Zn(0.01ml mp84317,0.04ml mp84318,0.0044ml m5819-10ml sample)	Jaswal	OK
70	Q1246-09A	A6342A	PS	02/02/25 18:10	PS For Se(0.01ml mp84317-10ml sample)	Jaswal	OK
71	CCV004	CCV004	CCV	02/02/25 18:13		Jaswal	OK
72	CCB004	CCB004	CCB	02/02/25 18:16		Jaswal	OK