

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Concentration Units: $\mu\text{g/L}$

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV001				ID: CCV001				ID: CCV002		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Antimony	200	210	104	2	500	490	98	4	500	99	1
Arsenic	200	210	106	1	500	510	103	1	530	106	3
Barium	100	100	103	1	2500	2500	100	3	2500	99	1
Beryllium	100	110	106	1	500	490	97	1	490	98	2
Cadmium	100	110	107	2	500	490	97	3	490	97	1
Chromium	100	100	103	1	500	510	102	0	520	104	3
Cobalt	100	100	103	0	500	510	103	0	530	107	3
Copper	100	110	107	1	5000	5000	100	1	5200	105	3
Lead	200	190	97	1	2500	2500	101	3	2500	101	1
Manganese	100	100	102	0	5000	5000	100	1	5200	103	3
Nickel	110	110	98	1	500	500	100	1	520	105	3
Selenium	200	210	106	2	500	490	98	3	490	98	0
Silver	50.0	51	102	2	500	480	97	3	480	96	1
Thallium	210	200	94	0	500	510	102	3	510	101	1
Vanadium	100	99	99	1	500	510	102	1	530	105	2
Zinc	200	200	102	1	5000	4900	98	1	5300	105	3

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Concentration Units: $\mu\text{g/L}$

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV003				ID: CCV004		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Antimony					500	490	98	1	490	98	2
Arsenic					500	500	101	1	510	103	0
Barium					2500	2500	99	1	2400	98	1
Beryllium					500	500	99	0	490	98	1
Cadmium					500	480	96	1	480	97	1
Chromium					500	500	101	1	510	101	0
Cobalt					500	520	104	2	520	105	1
Copper					5000	5100	102	1	5100	102	0
Lead					2500	2500	101	1	2500	101	0
Manganese					5000	5000	100	1	5000	100	1
Nickel					500	500	101	1	510	102	1
Selenium					500	490	98	1	500	99	1
Silver					500	480	96	1	500	100	1
Thallium					500	510	102	1	510	101	1
Vanadium					500	510	102	1	510	102	1
Zinc					5000	5100	102	1	5100	101	1