

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1249 SAS No.: Q1249 SDG NO.: Q1249

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 02/06/2025

Lab File ID: BN036320.D Time Analyzed: 05:40

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	4298	17.149	3803	21.34	4157	23.627
UPPER LIMIT	8596	17.649	7606	21.84	8314	24.127
LOWER LIMIT	2149	16.649	1901.5	20.84	2078.5	23.127
EPA SAMPLE NO.						
01 PB166470BSD	4578	17.15	4182	21.34	4261	23.63

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.