

Surrogate Summary

SDG No.: Q1249

Client: Tetra Tech NUS, Inc.

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1249-01	BP-VPB-192-GW-460-462	1,2-Dichloroethane-d4	50	45.2	90	81	118
		Dibromofluoromethane	50	49.2	98	80	119
		Toluene-d8	50	49.4	99	89	112
		4-Bromofluorobenzene	50	45.6	91	85	114
VN0204WBL01	VN0204WBL01	1,2-Dichloroethane-d4	50	49.9	100	81	118
		Dibromofluoromethane	50	50.3	101	80	119
		Toluene-d8	50	49.2	98	89	112
		4-Bromofluorobenzene	50	44.2	88	85	114
VN0204WBS02	VN0204WBS02	1,2-Dichloroethane-d4	50	45.2	90	81	118
		Dibromofluoromethane	50	48.7	97	80	119
		Toluene-d8	50	49.8	100	89	112
		4-Bromofluorobenzene	50	47.8	96	85	114
VN0204WBSD0	VN0204WBSD02	1,2-Dichloroethane-d4	50	42.7	85	81	118
		Dibromofluoromethane	50	52.2	104	80	119
		Toluene-d8	50	48.0	96	89	112
		4-Bromofluorobenzene	50	47.9	96	85	114