

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1249

Client: Tetra Tech NUS, Inc.

Analytical Method: SW8260-Low

Datafile : VN085651.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0204WBS02	Chloromethane	20	20.6	ug/L	103			50	139	
	Vinyl chloride	20	25.2	ug/L	126			58	137	
	Bromomethane	20	24.8	ug/L	124			53	141	
	Chloroethane	20	24.7	ug/L	124			60	138	
	Trichlorofluoromethane	20	20.6	ug/L	103			65	141	
	1,1,2-Trichlorotrifluoroethane	20	21.5	ug/L	108			70	136	
	1,1-Dichloroethene	20	21.6	ug/L	108			71	131	
	Acetone	100	93.0	ug/L	93			39	160	
	Carbon disulfide	20	21.6	ug/L	108			64	133	
	Methyl tert-butyl Ether	20	20.3	ug/L	102			71	124	
	Methylene Chloride	20	21.3	ug/L	106			74	124	
	trans-1,2-Dichloroethene	20	21.6	ug/L	108			75	124	
	1,1-Dichloroethane	20	19.8	ug/L	99			77	125	
	2-Butanone	100	99.4	ug/L	99			56	143	
	Carbon Tetrachloride	20	19.1	ug/L	96			72	136	
	cis-1,2-Dichloroethene	20	21.0	ug/L	105			78	123	
	Chloroform	20	20.1	ug/L	101			79	124	
	1,1,1-Trichloroethane	20	20.0	ug/L	100			74	131	
	Methylcyclohexane	20	20.1	ug/L	101			72	132	
	Benzene	20	20.6	ug/L	103			79	120	
	1,2-Dichloroethane	20	18.2	ug/L	91			73	128	
	Trichloroethene	20	19.8	ug/L	99			79	123	
	1,2-Dichloropropane	20	19.7	ug/L	99			78	122	
	Bromodichloromethane	20	19.3	ug/L	97			79	125	
	4-Methyl-2-Pentanone	100	98.0	ug/L	98			67	130	
	Toluene	20	21.7	ug/L	109			80	121	
	t-1,3-Dichloropropene	20	19.6	ug/L	98			73	127	
	cis-1,3-Dichloropropene	20	20.4	ug/L	102			75	124	
	1,1,2-Trichloroethane	20	20.9	ug/L	104			80	119	
	2-Hexanone	100	100	ug/L	100			57	139	
	Dibromochloromethane	20	19.9	ug/L	100			74	126	
	Tetrachloroethene	20	21.2	ug/L	106			74	129	
	Chlorobenzene	20	20.7	ug/L	104			82	118	
	Ethyl Benzene	20	20.6	ug/L	103			79	121	
	m/p-Xylenes	40	43.4	ug/L	109			80	121	
	o-Xylene	20	21.1	ug/L	106			78	122	
	Styrene	20	21.9	ug/L	110			78	123	
	Bromoform	20	21.2	ug/L	106			66	130	
	Isopropylbenzene	20	20.8	ug/L	104			72	131	
	1,1,2,2-Tetrachloroethane	20	21.0	ug/L	105			71	121	
	1,3-Dichlorobenzene	20	19.9	ug/L	100			80	119	
	1,4-Dichlorobenzene	20	19.4	ug/L	97			79	118	
	1,2-Dichlorobenzene	20	19.6	ug/L	98			80	119	