

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q1249 SAS No.: Q1249 SDG NO.: Q1249
Lab File ID: VN085644.D BFB Injection Date: 02/04/2025
Instrument ID: MSVOA_N BFB Injection Time: 09:55
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	19.4
75	30.0 - 60.0% of mass 95	51.3
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.3
173	Less than 2.0% of mass 174	1.2 (1.4) 1
174	50.0 - 100.0% of mass 95	80.1
175	5.0 - 9.0% of mass 174	5.8 (7.2) 1
176	95.0 - 101.0% of mass 174	77.2 (96.3) 1
177	5.0 - 9.0% of mass 176	4.4 (5.7) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VN085645.D	02/04/2025	11:05
VN0204WBL01	VN0204WBL01	VN085647.D	02/04/2025	12:31
VN0204WBS02	VN0204WBS02	VN085651.D	02/04/2025	14:51
VN0204WBSD02	VN0204WBSD02	VN085655.D	02/04/2025	16:37
BP-VPB-192-GW-460-462	Q1249-01	VN085661.D	02/04/2025	19:02
VSTDCCC050EC	VSTDCCC050	VN085664.D	02/04/2025	20:15