

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1250 SAS No.: Q1250 SDG NO.: Q1250

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 02/06/2025

Lab File ID: BN036320.D Time Analyzed: 05:40

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	2017	7.775	4328	10.56	2158	14.41
UPPER LIMIT	4034	8.275	8656	11.062	4316	14.909
LOWER LIMIT	1008.5	7.275	2164	10.062	1079	13.909
EPA SAMPLE NO.						
01 PB166470BSD	2053	7.78	4504	10.56	2275	14.41

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.