

LAB CHRONICLE

OrderID:	Q1250	OrderDate:	1/31/2025 10:38:00 AM
Client:	Tetra Tech NUS, Inc.	Project:	CTO WE13
Contact:	Ernie Wu	Location:	N31,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1250-01	BP-VPB-192-TB-2025 0127	Water			01/27/25			01/30/25
			VOCMS Group1	8260-Low			02/10/25	
Q1250-02	BP-VPB-192-GW-420- 422	Water			01/29/25			01/30/25
			VOCMS Group1	8260-Low			02/10/25	
Q1250-03	BP-VPB-192-GW-300- 302	Water			01/27/25			01/30/25
			VOCMS Group1	8260-Low			02/10/25	
Q1250-04	BP-VPB-192-GW-320- 322	Water			01/27/25			01/30/25
			VOCMS Group1	8260-Low			02/10/25	
Q1250-05	BP-VPB-192-GW-340- 342	Water			01/27/25			01/30/25
			VOCMS Group1	8260-Low			02/10/25	
Q1250-06	BP-VPB-192-GW-360- 362	Water			01/28/25			01/30/25
			VOCMS Group1	8260-Low			02/10/25	
Q1250-10	BP-VPB-192-GW-380- 382	Water			01/28/25			01/30/25
			VOCMS Group1	8260-Low			02/10/25	
Q1250-11	BP-VPB-192-GW-400- 402	Water			01/28/25			01/30/25
			VOCMS Group1	8260-Low			02/10/25	
Q1250-12	BP-VPB-192-GW-440- 442	Water			01/29/25			01/30/25
			VOCMS Group1	8260-Low			02/10/25	

Hit Summary Sheet
SW-846

SDG No.: Q1250
Client: Tetra Tech NUS, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID: Q1250-01	BP-VPB-192-TB-20250127 BP-VPB-192-TB-20 Water	Acetone		2.90	J	1.40	3.80	5.00	ug/L
		Total Voc :		2.90					
		Total Concentration:		2.90					
Client ID: Q1250-02	BP-VPB-192-GW-420-422 BP-VPB-192-GW-4 Water	Acetone		4.80	J	1.40	3.80	5.00	ug/L
Q1250-02	BP-VPB-192-GW-4 Water	Carbon Disulfide		0.54	J	0.32	0.75	1.00	ug/L
		Total Voc :		5.34					
		Total Concentration:		5.34					
Client ID: Q1250-03	BP-VPB-192-GW-300-302 BP-VPB-192-GW-3 Water	Acetone		6.00		1.40	3.80	5.00	ug/L
Q1250-03	BP-VPB-192-GW-3 Water	Carbon Disulfide		0.36	J	0.32	0.75	1.00	ug/L
		Total Voc :		6.36					
		Total Concentration:		6.36					
Client ID: Q1250-04	BP-VPB-192-GW-320-322 BP-VPB-192-GW-3 Water	Acetone		8.60		1.40	3.80	5.00	ug/L
Q1250-04	BP-VPB-192-GW-3 Water	Carbon Disulfide		0.51	J	0.32	0.75	1.00	ug/L
		Total Voc :		9.11					
		Total Concentration:		9.11					
Client ID: Q1250-05	BP-VPB-192-GW-340-342 BP-VPB-192-GW-3 Water	Acetone		8.70		1.40	3.80	5.00	ug/L
Q1250-05	BP-VPB-192-GW-3 Water	Carbon Disulfide		0.81	J	0.32	0.75	1.00	ug/L
		Total Voc :		9.51					
		Total Concentration:		9.51					
Client ID: Q1250-06	BP-VPB-192-GW-360-362 BP-VPB-192-GW-3 Water	Acetone		4.80	J	1.40	3.80	5.00	ug/L
		Total Voc :		4.80					
		Total Concentration:		4.80					
Client ID: Q1250-10	BP-VPB-192-GW-380-382 BP-VPB-192-GW-3 Water	Acetone		7.50		1.40	3.80	5.00	ug/L
Q1250-10	BP-VPB-192-GW-3 Water	Carbon Disulfide		0.86	J	0.32	0.75	1.00	ug/L
		Total Voc :		8.36					
		Total Concentration:		8.36					
Client ID: Q1250-11	BP-VPB-192-GW-400-402 BP-VPB-192-GW-4 Water	Acetone		12.8		1.40	3.80	5.00	ug/L
Q1250-11	BP-VPB-192-GW-4 Water	Carbon Disulfide		0.34	J	0.32	0.75	1.00	ug/L
		Total Voc :		13.1					
		Total Concentration:		13.1					
Client ID: Q1250-12	BP-VPB-192-GW-440-442 BP-VPB-192-GW-4 Water	Acetone		11.0		1.40	3.80	5.00	ug/L
Q1250-12	BP-VPB-192-GW-4 Water	Carbon Disulfide		0.45	J	0.32	0.75	1.00	ug/L

Hit Summary Sheet
SW-846

SDG No.: Q1250

Client: Tetra Tech NUS, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	LOD	RDL	Units
Total Voc :				11.4					
Total Concentration:				11.4					



SAMPLE DATA

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	01/27/25	
Project:	CTO WE13		Date Received:	01/30/25	
Client Sample ID:	BP-VPB-192-TB-20250127		SDG No.:	Q1250	
Lab Sample ID:	Q1250-01		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044881.D	1		02/10/25 15:41	VX021025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	0.50	U	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.26	0.75	1.00	ug/L
67-64-1	Acetone	2.90	J	1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.25	0.75	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	01/27/25	
Project:	CTO WE13		Date Received:	01/30/25	
Client Sample ID:	BP-VPB-192-TB-20250127		SDG No.:	Q1250	
Lab Sample ID:	Q1250-01		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044881.D	1		02/10/25 15:41	VX021025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	53.0		81 - 118		106%	SPK: 50
1868-53-7	Dibromofluoromethane	49.5		80 - 119		99%	SPK: 50
2037-26-5	Toluene-d8	50.2		89 - 112		100%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.6		85 - 114		101%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	92700	5.544				
540-36-3	1,4-Difluorobenzene	190000	6.757				
3114-55-4	Chlorobenzene-d5	171000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	69200	12.018				
TENTATIVE IDENTIFIED COMPOUNDS							
75-43-4	Dichlorofluoromethane	N.D					

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	01/27/25	
Project:	CTO WE13		Date Received:	01/30/25	
Client Sample ID:	BP-VPB-192-TB-20250127		SDG No.:	Q1250	
Lab Sample ID:	Q1250-01		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044881.D	1		02/10/25 15:41	VX021025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	01/29/25	
Project:	CTO WE13		Date Received:	01/30/25	
Client Sample ID:	BP-VPB-192-GW-420-422		SDG No.:	Q1250	
Lab Sample ID:	Q1250-02		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044880.D	1		02/10/25 15:18	VX021025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	0.50	U	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.26	0.75	1.00	ug/L
67-64-1	Acetone	4.80	J	1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.54	J	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.25	0.75	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	01/29/25	
Project:	CTO WE13		Date Received:	01/30/25	
Client Sample ID:	BP-VPB-192-GW-420-422		SDG No.:	Q1250	
Lab Sample ID:	Q1250-02		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044880.D	1		02/10/25 15:18	VX021025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	53.1		81 - 118		106%	SPK: 50
1868-53-7	Dibromofluoromethane	51.2		80 - 119		102%	SPK: 50
2037-26-5	Toluene-d8	50.9		89 - 112		102%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.7		85 - 114		103%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	96900	5.549				
540-36-3	1,4-Difluorobenzene	197000	6.756				
3114-55-4	Chlorobenzene-d5	180000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	78400	12.018				
TENTATIVE IDENTIFIED COMPOUNDS							
75-43-4	Dichlorofluoromethane	N.D					

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	01/29/25	
Project:	CTO WE13		Date Received:	01/30/25	
Client Sample ID:	BP-VPB-192-GW-420-422		SDG No.:	Q1250	
Lab Sample ID:	Q1250-02		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044880.D	1		02/10/25 15:18	VX021025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	01/27/25	
Project:	CTO WE13		Date Received:	01/30/25	
Client Sample ID:	BP-VPB-192-GW-300-302		SDG No.:	Q1250	
Lab Sample ID:	Q1250-03		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044882.D	1		02/10/25 16:04	VX021025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	0.50	U	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.26	0.75	1.00	ug/L
67-64-1	Acetone	6.00		1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.36	J	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.25	0.75	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	01/27/25	
Project:	CTO WE13		Date Received:	01/30/25	
Client Sample ID:	BP-VPB-192-GW-300-302		SDG No.:	Q1250	
Lab Sample ID:	Q1250-03		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044882.D	1		02/10/25 16:04	VX021025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	53.2		81 - 118		106%	SPK: 50
1868-53-7	Dibromofluoromethane	51.1		80 - 119		102%	SPK: 50
2037-26-5	Toluene-d8	50.5		89 - 112		101%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.8		85 - 114		104%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	99300	5.55				
540-36-3	1,4-Difluorobenzene	201000	6.757				
3114-55-4	Chlorobenzene-d5	184000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	78900	12.018				
TENTATIVE IDENTIFIED COMPOUNDS							
75-43-4	Dichlorofluoromethane	N.D					

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	01/27/25	
Project:	CTO WE13		Date Received:	01/30/25	
Client Sample ID:	BP-VPB-192-GW-300-302		SDG No.:	Q1250	
Lab Sample ID:	Q1250-03		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044882.D	1		02/10/25 16:04	VX021025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	01/27/25	
Project:	CTO WE13		Date Received:	01/30/25	
Client Sample ID:	BP-VPB-192-GW-320-322		SDG No.:	Q1250	
Lab Sample ID:	Q1250-04		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044883.D	1		02/10/25 16:27	VX021025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	0.50	U	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.26	0.75	1.00	ug/L
67-64-1	Acetone	8.60		1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.51	J	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.25	0.75	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	01/27/25	
Project:	CTO WE13		Date Received:	01/30/25	
Client Sample ID:	BP-VPB-192-GW-320-322		SDG No.:	Q1250	
Lab Sample ID:	Q1250-04		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044883.D	1		02/10/25 16:27	VX021025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	52.3		81 - 118		105%	SPK: 50
1868-53-7	Dibromofluoromethane	49.7		80 - 119		99%	SPK: 50
2037-26-5	Toluene-d8	50.5		89 - 112		101%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.9		85 - 114		104%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	96100	5.55				
540-36-3	1,4-Difluorobenzene	195000	6.757				
3114-55-4	Chlorobenzene-d5	175000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	76700	12.018				
TENTATIVE IDENTIFIED COMPOUNDS							
75-43-4	Dichlorofluoromethane	N.D					

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	01/27/25	
Project:	CTO WE13		Date Received:	01/30/25	
Client Sample ID:	BP-VPB-192-GW-320-322		SDG No.:	Q1250	
Lab Sample ID:	Q1250-04		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044883.D	1		02/10/25 16:27	VX021025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	01/27/25	
Project:	CTO WE13		Date Received:	01/30/25	
Client Sample ID:	BP-VPB-192-GW-340-342		SDG No.:	Q1250	
Lab Sample ID:	Q1250-05		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044884.D	1		02/10/25 16:51	VX021025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	0.50	U	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.26	0.75	1.00	ug/L
67-64-1	Acetone	8.70		1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.81	J	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.25	0.75	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	01/27/25	
Project:	CTO WE13		Date Received:	01/30/25	
Client Sample ID:	BP-VPB-192-GW-340-342		SDG No.:	Q1250	
Lab Sample ID:	Q1250-05		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044884.D	1		02/10/25 16:51	VX021025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	52.3		81 - 118		105%	SPK: 50
1868-53-7	Dibromofluoromethane	51.1		80 - 119		102%	SPK: 50
2037-26-5	Toluene-d8	49.7		89 - 112		99%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.4		85 - 114		105%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	97000	5.544				
540-36-3	1,4-Difluorobenzene	194000	6.757				
3114-55-4	Chlorobenzene-d5	176000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	75900	12.018				
TENTATIVE IDENTIFIED COMPOUNDS							
75-43-4	Dichlorofluoromethane	N.D					

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	01/27/25	
Project:	CTO WE13		Date Received:	01/30/25	
Client Sample ID:	BP-VPB-192-GW-340-342		SDG No.:	Q1250	
Lab Sample ID:	Q1250-05		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044884.D	1		02/10/25 16:51	VX021025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	01/28/25	
Project:	CTO WE13		Date Received:	01/30/25	
Client Sample ID:	BP-VPB-192-GW-360-362		SDG No.:	Q1250	
Lab Sample ID:	Q1250-06		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044888.D	1		02/10/25 18:23	VX021025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	0.50	U	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.26	0.75	1.00	ug/L
67-64-1	Acetone	4.80	J	1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.25	0.75	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	01/28/25	
Project:	CTO WE13		Date Received:	01/30/25	
Client Sample ID:	BP-VPB-192-GW-360-362		SDG No.:	Q1250	
Lab Sample ID:	Q1250-06		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044888.D	1		02/10/25 18:23	VX021025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	52.6		81 - 118		105%	SPK: 50
1868-53-7	Dibromofluoromethane	51.2		80 - 119		102%	SPK: 50
2037-26-5	Toluene-d8	51.0		89 - 112		102%	SPK: 50
460-00-4	4-Bromofluorobenzene	53.2		85 - 114		106%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	93900	5.543				
540-36-3	1,4-Difluorobenzene	188000	6.757				
3114-55-4	Chlorobenzene-d5	174000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	73200	12.018				
TENTATIVE IDENTIFIED COMPOUNDS							
75-43-4	Dichlorofluoromethane	N.D					

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	01/28/25	
Project:	CTO WE13		Date Received:	01/30/25	
Client Sample ID:	BP-VPB-192-GW-360-362		SDG No.:	Q1250	
Lab Sample ID:	Q1250-06		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044888.D	1		02/10/25 18:23	VX021025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	01/28/25	
Project:	CTO WE13		Date Received:	01/30/25	
Client Sample ID:	BP-VPB-192-GW-380-382		SDG No.:	Q1250	
Lab Sample ID:	Q1250-10		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044885.D	1		02/10/25 17:14	VX021025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	0.50	U	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.26	0.75	1.00	ug/L
67-64-1	Acetone	7.50		1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.86	J	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.25	0.75	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	01/28/25	
Project:	CTO WE13		Date Received:	01/30/25	
Client Sample ID:	BP-VPB-192-GW-380-382		SDG No.:	Q1250	
Lab Sample ID:	Q1250-10		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044885.D	1		02/10/25 17:14	VX021025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	53.8		81 - 118		108%	SPK: 50
1868-53-7	Dibromofluoromethane	51.0		80 - 119		102%	SPK: 50
2037-26-5	Toluene-d8	50.9		89 - 112		102%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.9		85 - 114		104%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	98000	5.544				
540-36-3	1,4-Difluorobenzene	197000	6.757				
3114-55-4	Chlorobenzene-d5	182000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	77300	12.018				
TENTATIVE IDENTIFIED COMPOUNDS							
75-43-4	Dichlorofluoromethane	N.D					

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	01/28/25	
Project:	CTO WE13		Date Received:	01/30/25	
Client Sample ID:	BP-VPB-192-GW-380-382		SDG No.:	Q1250	
Lab Sample ID:	Q1250-10		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044885.D	1		02/10/25 17:14	VX021025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	01/28/25	
Project:	CTO WE13		Date Received:	01/30/25	
Client Sample ID:	BP-VPB-192-GW-400-402		SDG No.:	Q1250	
Lab Sample ID:	Q1250-11		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044886.D	1		02/10/25 17:37	VX021025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	0.50	U	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.26	0.75	1.00	ug/L
67-64-1	Acetone	12.8		1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.34	J	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.25	0.75	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	01/28/25	
Project:	CTO WE13		Date Received:	01/30/25	
Client Sample ID:	BP-VPB-192-GW-400-402		SDG No.:	Q1250	
Lab Sample ID:	Q1250-11		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044886.D	1		02/10/25 17:37	VX021025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	53.8		81 - 118		108%	SPK: 50
1868-53-7	Dibromofluoromethane	51.3		80 - 119		103%	SPK: 50
2037-26-5	Toluene-d8	51.0		89 - 112		102%	SPK: 50
460-00-4	4-Bromofluorobenzene	53.5		85 - 114		107%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	94400	5.549				
540-36-3	1,4-Difluorobenzene	192000	6.756				
3114-55-4	Chlorobenzene-d5	179000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	76700	12.018				
TENTATIVE IDENTIFIED COMPOUNDS							
75-43-4	Dichlorofluoromethane	N.D					

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	01/28/25	
Project:	CTO WE13		Date Received:	01/30/25	
Client Sample ID:	BP-VPB-192-GW-400-402		SDG No.:	Q1250	
Lab Sample ID:	Q1250-11		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044886.D	1		02/10/25 17:37	VX021025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	01/29/25	
Project:	CTO WE13		Date Received:	01/30/25	
Client Sample ID:	BP-VPB-192-GW-440-442		SDG No.:	Q1250	
Lab Sample ID:	Q1250-12		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044887.D	1		02/10/25 18:00	VX021025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	0.50	U	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.26	0.75	1.00	ug/L
67-64-1	Acetone	11.0		1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.45	J	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.25	0.75	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	01/29/25	
Project:	CTO WE13		Date Received:	01/30/25	
Client Sample ID:	BP-VPB-192-GW-440-442		SDG No.:	Q1250	
Lab Sample ID:	Q1250-12		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044887.D	1		02/10/25 18:00	VX021025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	52.9		81 - 118		106%	SPK: 50
1868-53-7	Dibromofluoromethane	51.0		80 - 119		102%	SPK: 50
2037-26-5	Toluene-d8	50.2		89 - 112		100%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.0		85 - 114		102%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	91100	5.544				
540-36-3	1,4-Difluorobenzene	183000	6.757				
3114-55-4	Chlorobenzene-d5	167000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	68200	12.018				
TENTATIVE IDENTIFIED COMPOUNDS							
75-43-4	Dichlorofluoromethane	N.D					

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	01/29/25	
Project:	CTO WE13		Date Received:	01/30/25	
Client Sample ID:	BP-VPB-192-GW-440-442		SDG No.:	Q1250	
Lab Sample ID:	Q1250-12		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044887.D	1		02/10/25 18:00	VX021025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



QC SUMMARY

Surrogate Summary

SDG No.: **Q1250**

Client: **Tetra Tech NUS, Inc.**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1250-01	BP-VPB-192-TB-20250127	1,2-Dichloroethane-d4	50	53.0	106	81	118
		Dibromofluoromethane	50	49.5	99	80	119
		Toluene-d8	50	50.2	100	89	112
		4-Bromofluorobenzene	50	50.6	101	85	114
Q1250-02	BP-VPB-192-GW-420-422	1,2-Dichloroethane-d4	50	53.1	106	81	118
		Dibromofluoromethane	50	51.2	102	80	119
		Toluene-d8	50	50.9	102	89	112
		4-Bromofluorobenzene	50	51.7	103	85	114
Q1250-03	BP-VPB-192-GW-300-302	1,2-Dichloroethane-d4	50	53.2	106	81	118
		Dibromofluoromethane	50	51.1	102	80	119
		Toluene-d8	50	50.5	101	89	112
		4-Bromofluorobenzene	50	51.8	104	85	114
Q1250-04	BP-VPB-192-GW-320-322	1,2-Dichloroethane-d4	50	52.3	105	81	118
		Dibromofluoromethane	50	49.7	99	80	119
		Toluene-d8	50	50.5	101	89	112
		4-Bromofluorobenzene	50	51.9	104	85	114
Q1250-05	BP-VPB-192-GW-340-342	1,2-Dichloroethane-d4	50	52.3	105	81	118
		Dibromofluoromethane	50	51.1	102	80	119
		Toluene-d8	50	49.7	99	89	112
		4-Bromofluorobenzene	50	52.4	105	85	114
Q1250-06	BP-VPB-192-GW-360-362	1,2-Dichloroethane-d4	50	52.6	105	81	118
		Dibromofluoromethane	50	51.2	102	80	119
		Toluene-d8	50	51.0	102	89	112
		4-Bromofluorobenzene	50	53.1	106	85	114
Q1250-07MS	BP-VPB-192-GW-360-362MS	1,2-Dichloroethane-d4	50	50.8	102	81	118
		Dibromofluoromethane	50	49.0	98	80	119
		Toluene-d8	50	49.5	99	89	112
		4-Bromofluorobenzene	50	49.3	99	85	114
Q1250-08MSD	BP-VPB-192-GW-360-362MSD	1,2-Dichloroethane-d4	50	50.2	100	81	118
		Dibromofluoromethane	50	50.4	101	80	119
		Toluene-d8	50	51.0	102	89	112
		4-Bromofluorobenzene	50	51.6	103	85	114
Q1250-10	BP-VPB-192-GW-380-382	1,2-Dichloroethane-d4	50	53.8	108	81	118
		Dibromofluoromethane	50	51.0	102	80	119
		Toluene-d8	50	50.9	102	89	112
		4-Bromofluorobenzene	50	52.0	104	85	114
Q1250-11	BP-VPB-192-GW-400-402	1,2-Dichloroethane-d4	50	53.9	108	81	118
		Dibromofluoromethane	50	51.3	103	80	119
		Toluene-d8	50	51.0	102	89	112
		4-Bromofluorobenzene	50	53.5	107	85	114
Q1250-12	BP-VPB-192-GW-440-442	1,2-Dichloroethane-d4	50	52.9	106	81	118
		Dibromofluoromethane	50	51.0	102	80	119
		Toluene-d8	50	50.2	100	89	112
		4-Bromofluorobenzene	50	51.0	102	85	114
VX0210WBL01	VX0210WBL01	1,2-Dichloroethane-d4	50	55.5	111	81	118
		Dibromofluoromethane	50	49.8	100	80	119
		Toluene-d8	50	50.5	101	89	112
		4-Bromofluorobenzene	50	51.2	102	85	114
VX0210WBS01	VX0210WBS01	1,2-Dichloroethane-d4	50	46.8	94	81	118
		Dibromofluoromethane	50	46.6	93	80	119

Surrogate Summary

SDG No.: Q1250

Client: Tetra Tech NUS, Inc.

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
VX0210WBS01	VX0210WBS01	Toluene-d8	50	47.0	94	89	112
		4-Bromofluorobenzene	50	47.2	94	85	114

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1250

Client: Tetra Tech NUS, Inc.

Analytical Method: SW8260-Low

Parameter	Spike	Sample	Result	Units	Rec		RPD		Limits		RPD
		Result			Qual	RPD	Qual	Low	High		
Lab Sample ID :	Q1250-07MS	Client Sample ID :	BP-VPB-192-GW-360-362MS					Datafile :	VX044889.D		
Chloromethane	50	0	46.2	ug/L	92				50	139	
Vinyl chloride	50	0	45.7	ug/L	91				58	137	
Bromomethane	50	0	47.6	ug/L	95				53	141	
Chloroethane	50	0	52.8	ug/L	106				60	138	
Trichlorofluoromethane	50	0	47.0	ug/L	94				65	141	
1,1,2-Trichlorotrifluoroethane	50	0	47.3	ug/L	95				70	136	
1,1-Dichloroethene	50	0	46.3	ug/L	93				71	131	
Acetone	250	4.80	250	ug/L	98				39	160	
Carbon disulfide	50	0	45.8	ug/L	92				64	133	
Methyl tert-butyl Ether	50	0	47.7	ug/L	95				71	124	
Methylene Chloride	50	0	47.0	ug/L	94				74	124	
trans-1,2-Dichloroethene	50	0	47.8	ug/L	96				75	124	
1,1-Dichloroethane	50	0	47.6	ug/L	95				77	125	
2-Butanone	250	0	250	ug/L	100				56	143	
Carbon Tetrachloride	50	0	45.0	ug/L	90				72	136	
cis-1,2-Dichloroethene	50	0	48.0	ug/L	96				78	123	
Chloroform	50	0	47.3	ug/L	95				79	124	
1,1,1-Trichloroethane	50	0	47.1	ug/L	94				74	131	
Methylcyclohexane	50	0	45.4	ug/L	91				72	132	
Benzene	50	0	46.4	ug/L	93				79	120	
1,2-Dichloroethane	50	0	47.6	ug/L	95				73	128	
Trichloroethene	50	0	45.6	ug/L	91				79	123	
1,2-Dichloropropane	50	0	46.7	ug/L	93				78	122	
Bromodichloromethane	50	0	47.7	ug/L	95				79	125	
4-Methyl-2-Pentanone	250	0	250	ug/L	100				67	130	
Toluene	50	0	47.4	ug/L	95				80	121	
t-1,3-Dichloropropene	50	0	47.0	ug/L	94				73	127	
cis-1,3-Dichloropropene	50	0	47.0	ug/L	94				75	124	
1,1,2-Trichloroethane	50	0	48.6	ug/L	97				80	119	
2-Hexanone	250	0	260	ug/L	104				57	139	
Dibromochloromethane	50	0	47.6	ug/L	95				74	126	
Tetrachloroethene	50	0	45.5	ug/L	91				74	129	
Chlorobenzene	50	0	46.6	ug/L	93				82	118	
Ethyl Benzene	50	0	47.0	ug/L	94				79	121	
m/p-Xylenes	100	0	93.5	ug/L	94				80	121	
o-Xylene	50	0	46.7	ug/L	93				78	122	
Styrene	50	0	47.9	ug/L	96				78	123	
Bromoform	50	0	47.4	ug/L	95				66	130	
Isopropylbenzene	50	0	47.0	ug/L	94				72	131	
1,1,2,2-Tetrachloroethane	50	0	47.1	ug/L	94				71	121	
1,3-Dichlorobenzene	50	0	46.0	ug/L	92				80	119	
1,4-Dichlorobenzene	50	0	45.2	ug/L	90				79	118	
1,2-Dichlorobenzene	50	0	47.5	ug/L	95				80	119	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1250

Client: Tetra Tech NUS, Inc.

Analytical Method: SW8260-Low

Limits											
Parameter	Spike	Sample	Result	Units	Rec		RPD	RPD		RPD	
		Result			Qual	Qual		Low	High		
Lab Sample ID :	Q1250-08MSD	Client Sample ID :	BP-VPB-192-GW-360-362MSD					Datafile :	VX044890.D		
Chloromethane	50	0	48.9	ug/L	98		6		50	139	20
Vinyl chloride	50	0	47.9	ug/L	96		5		58	137	20
Bromomethane	50	0	50.2	ug/L	100		5		53	141	20
Chloroethane	50	0	52.3	ug/L	105		1		60	138	20
Trichlorofluoromethane	50	0	49.5	ug/L	99		5		65	141	20
1,1,2-Trichlorotrifluoroethane	50	0	49.6	ug/L	99		5		70	136	20
1,1-Dichloroethene	50	0	50.2	ug/L	100		8		71	131	20
Acetone	250	4.80	260	ug/L	102		4		39	160	20
Carbon disulfide	50	0	49.3	ug/L	99		7		64	133	20
Methyl tert-butyl Ether	50	0	51.1	ug/L	102		7		71	124	20
Methylene Chloride	50	0	50.3	ug/L	101		7		74	124	20
trans-1,2-Dichloroethene	50	0	50.8	ug/L	102		6		75	124	20
1,1-Dichloroethane	50	0	51.2	ug/L	102		7		77	125	20
2-Butanone	250	0	270	ug/L	108		8		56	143	20
Carbon Tetrachloride	50	0	50.5	ug/L	101		12		72	136	20
cis-1,2-Dichloroethene	50	0	50.7	ug/L	101		5		78	123	20
Chloroform	50	0	50.8	ug/L	102		7		79	124	20
1,1,1-Trichloroethane	50	0	51.0	ug/L	102		8		74	131	20
Methylcyclohexane	50	0	49.1	ug/L	98		8		72	132	20
Benzene	50	0	51.3	ug/L	103		10		79	120	20
1,2-Dichloroethane	50	0	51.8	ug/L	104		8		73	128	20
Trichloroethene	50	0	51.2	ug/L	102		12		79	123	20
1,2-Dichloropropane	50	0	51.2	ug/L	102		9		78	122	20
Bromodichloromethane	50	0	52.4	ug/L	105		9		79	125	20
4-Methyl-2-Pentanone	250	0	280	ug/L	112		11		67	130	20
Toluene	50	0	51.8	ug/L	104		9		80	121	20
t-1,3-Dichloropropene	50	0	51.8	ug/L	104		10		73	127	20
cis-1,3-Dichloropropene	50	0	52.5	ug/L	105		11		75	124	20
1,1,2-Trichloroethane	50	0	52.6	ug/L	105		8		80	119	20
2-Hexanone	250	0	280	ug/L	112		7		57	139	20
Dibromochloromethane	50	0	52.8	ug/L	106		10		74	126	20
Tetrachloroethene	50	0	48.5	ug/L	97		6		74	129	20
Chlorobenzene	50	0	50.5	ug/L	101		8		82	118	20
Ethyl Benzene	50	0	50.4	ug/L	101		7		79	121	20
m/p-Xylenes	100	0	100	ug/L	100		7		80	121	20
o-Xylene	50	0	50.4	ug/L	101		8		78	122	20
Styrene	50	0	52.6	ug/L	105		9		78	123	20
Bromoform	50	0	53.2	ug/L	106		12		66	130	20
Isopropylbenzene	50	0	50.6	ug/L	101		7		72	131	20
1,1,2,2-Tetrachloroethane	50	0	50.6	ug/L	101		7		71	121	20
1,3-Dichlorobenzene	50	0	49.9	ug/L	100		8		80	119	20
1,4-Dichlorobenzene	50	0	49.2	ug/L	98		8		79	118	20
1,2-Dichlorobenzene	50	0	51.7	ug/L	103		8		80	119	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1250

Client: Tetra Tech NUS, Inc.

Analytical Method: SW8260-Low

Datafile : VX044878.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0210WBS01	Chloromethane	20	19.2	ug/L	96			50	139	
	Vinyl chloride	20	18.2	ug/L	91			58	137	
	Bromomethane	20	19.8	ug/L	99			53	141	
	Chloroethane	20	22.6	ug/L	113			60	138	
	Trichlorofluoromethane	20	19.4	ug/L	97			65	141	
	1,1,2-Trichlorotrifluoroethane	20	20.5	ug/L	103			70	136	
	1,1-Dichloroethene	20	18.8	ug/L	94			71	131	
	Acetone	100	110	ug/L	110			39	160	
	Carbon disulfide	20	19.5	ug/L	98			64	133	
	Methyl tert-butyl Ether	20	19.6	ug/L	98			71	124	
	Methylene Chloride	20	19.3	ug/L	97			74	124	
	trans-1,2-Dichloroethene	20	19.6	ug/L	98			75	124	
	1,1-Dichloroethane	20	19.8	ug/L	99			77	125	
	2-Butanone	100	110	ug/L	110			56	143	
	Carbon Tetrachloride	20	19.8	ug/L	99			72	136	
	cis-1,2-Dichloroethene	20	20.0	ug/L	100			78	123	
	Chloroform	20	19.7	ug/L	99			79	124	
	1,1,1-Trichloroethane	20	19.6	ug/L	98			74	131	
	Methylcyclohexane	20	21.2	ug/L	106			72	132	
	Benzene	20	20.1	ug/L	101			79	120	
	1,2-Dichloroethane	20	20.4	ug/L	102			73	128	
	Trichloroethene	20	19.8	ug/L	99			79	123	
	1,2-Dichloropropane	20	19.9	ug/L	100			78	122	
	Bromodichloromethane	20	20.1	ug/L	101			79	125	
	4-Methyl-2-Pentanone	100	110	ug/L	110			67	130	
	Toluene	20	20.1	ug/L	101			80	121	
	t-1,3-Dichloropropene	20	19.8	ug/L	99			73	127	
	cis-1,3-Dichloropropene	20	20.3	ug/L	102			75	124	
	1,1,2-Trichloroethane	20	19.9	ug/L	100			80	119	
	2-Hexanone	100	110	ug/L	110			57	139	
	Dibromochloromethane	20	19.9	ug/L	100			74	126	
	Tetrachloroethene	20	20.8	ug/L	104			74	129	
	Chlorobenzene	20	20.7	ug/L	104			82	118	
	Ethyl Benzene	20	20.5	ug/L	103			79	121	
	m/p-Xylenes	40	42.0	ug/L	105			80	121	
	o-Xylene	20	20.6	ug/L	103			78	122	
	Styrene	20	21.0	ug/L	105			78	123	
	Bromoform	20	20.5	ug/L	103			66	130	
	Isopropylbenzene	20	19.8	ug/L	99			72	131	
	1,1,2,2-Tetrachloroethane	20	19.9	ug/L	100			71	121	
	1,3-Dichlorobenzene	20	19.6	ug/L	98			80	119	
	1,4-Dichlorobenzene	20	19.9	ug/L	100			79	118	
	1,2-Dichlorobenzene	20	20.3	ug/L	102			80	119	

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VX0210WBL01

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM Case No.: Q1250

SAS No.: Q1250 SDG NO.: Q1250

Lab File ID: VX044877.D

Lab Sample ID: VX0210WBL01

Date Analyzed: 02/10/2025

Time Analyzed: 14:06

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_X

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VX0210WBS01	VX0210WBS01	VX044878.D	02/10/2025
BP-VPB-192-GW-420-422	Q1250-02	VX044880.D	02/10/2025
BP-VPB-192-TB-20250127	Q1250-01	VX044881.D	02/10/2025
BP-VPB-192-GW-300-302	Q1250-03	VX044882.D	02/10/2025
BP-VPB-192-GW-320-322	Q1250-04	VX044883.D	02/10/2025
BP-VPB-192-GW-340-342	Q1250-05	VX044884.D	02/10/2025
BP-VPB-192-GW-380-382	Q1250-10	VX044885.D	02/10/2025
BP-VPB-192-GW-400-402	Q1250-11	VX044886.D	02/10/2025
BP-VPB-192-GW-440-442	Q1250-12	VX044887.D	02/10/2025
BP-VPB-192-GW-360-362	Q1250-06	VX044888.D	02/10/2025
BP-VPB-192-GW-360-362MS	Q1250-07MS	VX044889.D	02/10/2025
BP-VPB-192-GW-360-362MSD	Q1250-08MSD	VX044890.D	02/10/2025

COMMENTS: _____

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q1250 SAS No.: Q1250 SDG NO.: Q1250
Lab File ID: VX044867.D BFB Injection Date: 02/10/2025
Instrument ID: MSVOA_X BFB Injection Time: 09:35
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	19.7
75	30.0 - 60.0% of mass 95	52.3
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.9
173	Less than 2.0% of mass 174	0.5 (0.6) 1
174	50.0 - 100.0% of mass 95	75.9
175	5.0 - 9.0% of mass 174	5.7 (7.5) 1
176	95.0 - 101.0% of mass 174	72.6 (95.7) 1
177	5.0 - 9.0% of mass 176	4.5 (6.2) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC001	VSTDIC001	VX044868.D	02/10/2025	10:25
VSTDIC005	VSTDIC005	VX044869.D	02/10/2025	10:48
VSTDIC020	VSTDIC020	VX044870.D	02/10/2025	11:11
VSTDIC050	VSTDIC050	VX044871.D	02/10/2025	11:34
VSTDIC100	VSTDIC100	VX044872.D	02/10/2025	12:05
VSTDIC150	VSTDIC150	VX044873.D	02/10/2025	12:28
VX0210WBL01	VX0210WBL01	VX044877.D	02/10/2025	14:06
VX0210WBS01	VX0210WBS01	VX044878.D	02/10/2025	14:29
BP-VPB-192-GW-420-422	Q1250-02	VX044880.D	02/10/2025	15:18
BP-VPB-192-TB-20250127	Q1250-01	VX044881.D	02/10/2025	15:41
BP-VPB-192-GW-300-302	Q1250-03	VX044882.D	02/10/2025	16:04
BP-VPB-192-GW-320-322	Q1250-04	VX044883.D	02/10/2025	16:27
BP-VPB-192-GW-340-342	Q1250-05	VX044884.D	02/10/2025	16:51
BP-VPB-192-GW-380-382	Q1250-10	VX044885.D	02/10/2025	17:14
BP-VPB-192-GW-400-402	Q1250-11	VX044886.D	02/10/2025	17:37
BP-VPB-192-GW-440-442	Q1250-12	VX044887.D	02/10/2025	18:00
BP-VPB-192-GW-360-362	Q1250-06	VX044888.D	02/10/2025	18:23
BP-VPB-192-GW-360-362MS	Q1250-07MS	VX044889.D	02/10/2025	18:46

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q1250 SAS No.: Q1250 SDG NO.: Q1250
Lab File ID: VX044867.D BFB Injection Date: 02/10/2025
Instrument ID: MSVOA_X BFB Injection Time: 09:35
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	19.7
75	30.0 - 60.0% of mass 95	52.3
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.9
173	Less than 2.0% of mass 174	0.5 (0.6) 1
174	50.0 - 100.0% of mass 95	75.9
175	5.0 - 9.0% of mass 174	5.7 (7.5) 1
176	95.0 - 101.0% of mass 174	72.6 (95.7) 1
177	5.0 - 9.0% of mass 176	4.5 (6.2) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
BP-VPB-192-GW-360-362MSD	Q1250-08MSD	VX044890.D	02/10/2025	19:09
VSTDCCC050EC	VSTDCCC050	VX044891.D	02/10/2025	19:32

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q1250 SAS No.: Q1250 SDG NO.: Q1250
Lab File ID: VX044871.D Date Analyzed: 02/10/2025
Instrument ID: MSVOA_X Time Analyzed: 11:34
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	127984	5.54	233736	6.76	204751	10.05
UPPER LIMIT	255968	6.043	467472	7.257	409502	10.549
LOWER LIMIT	63992	5.043	116868	6.257	102376	9.549
EPA SAMPLE NO.						
BP-VPB-192-TB-20250127	92731	5.54	189692	6.76	170630	10.05
BP-VPB-192-GW-420-422	96890	5.55	196670	6.76	180490	10.05
BP-VPB-192-GW-300-302	99257	5.55	200954	6.76	183816	10.05
BP-VPB-192-GW-320-322	96132	5.55	194678	6.76	175107	10.05
BP-VPB-192-GW-340-342	96963	5.54	193723	6.76	175745	10.05
BP-VPB-192-GW-360-362	93911	5.54	187948	6.76	173877	10.05
BP-VPB-192-GW-360-362MS	123128	5.55	231125	6.76	204212	10.05
BP-VPB-192-GW-360-362MSD	113791	5.54	209107	6.76	187126	10.05
BP-VPB-192-GW-380-382	98000	5.54	197272	6.76	182089	10.05
BP-VPB-192-GW-400-402	94362	5.55	191816	6.76	178953	10.05
BP-VPB-192-GW-440-442	91120	5.54	183168	6.76	166878	10.05
VX0210WBL01	95840	5.54	200206	6.76	182254	10.05
VX0210WBS01	143675	5.54	262165	6.76	224717	10.05

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q1250 SAS No.: Q1250 SDG NO.: Q1250
Lab File ID: VX044871.D Date Analyzed: 02/10/2025
Instrument ID: MSVOA_X Time Analyzed: 11:34
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	92648	12.018				
UPPER LIMIT	185296	12.518				
LOWER LIMIT	46324	11.518				
EPA SAMPLE NO.						
BP-VPB-192-TB-20250127	69234	12.02				
BP-VPB-192-GW-420-422	78413	12.02				
BP-VPB-192-GW-300-302	78851	12.02				
BP-VPB-192-GW-320-322	76744	12.02				
BP-VPB-192-GW-340-342	75903	12.02				
BP-VPB-192-GW-360-362	73228	12.02				
BP-VPB-192-GW-360-362MS	88501	12.02				
BP-VPB-192-GW-360-362MSD	83002	12.02				
BP-VPB-192-GW-380-382	77317	12.02				
BP-VPB-192-GW-400-402	76716	12.02				
BP-VPB-192-GW-440-442	68152	12.02				
VX0210WBL01	76715	12.02				
VX0210WBS01	103194	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	CTO WE13		Date Received:	
Client Sample ID:	VX0210WBL01		SDG No.:	Q1250
Lab Sample ID:	VX0210WBL01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044877.D	1		02/10/25 14:06	VX021025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	0.50	U	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.26	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.25	0.75	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	CTO WE13		Date Received:	
Client Sample ID:	VX0210WBL01		SDG No.:	Q1250
Lab Sample ID:	VX0210WBL01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044877.D	1		02/10/25 14:06	VX021025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	55.5		81 - 118		111%	SPK: 50
1868-53-7	Dibromofluoromethane	49.8		80 - 119		100%	SPK: 50
2037-26-5	Toluene-d8	50.5		89 - 112		101%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.2		85 - 114		102%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	95800	5.544				
540-36-3	1,4-Difluorobenzene	200000	6.757				
3114-55-4	Chlorobenzene-d5	182000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	76700	12.018				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	CTO WE13		Date Received:	
Client Sample ID:	VX0210WBS01		SDG No.:	Q1250
Lab Sample ID:	VX0210WBS01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044878.D	1		02/10/25 14:29	VX021025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	19.2		0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	18.2		0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	19.8		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	22.6		0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	19.4		0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	20.5		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	18.8		0.26	0.75	1.00	ug/L
67-64-1	Acetone	110		1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	19.5		0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	19.6		0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	19.3		0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	19.6		0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	19.8		0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	110		1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	19.8		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	20.0		0.25	0.75	1.00	ug/L
67-66-3	Chloroform	19.7		0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	19.6		0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	21.2		0.19	0.50	1.00	ug/L
71-43-2	Benzene	20.1		0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	20.4		0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	19.8		0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	19.9		0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	20.1		0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	110		0.75	2.50	5.00	ug/L
108-88-3	Toluene	20.1		0.18	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	19.8		0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	20.3		0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	19.9		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	110		1.10	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	CTO WE13		Date Received:	
Client Sample ID:	VX0210WBS01		SDG No.:	Q1250
Lab Sample ID:	VX0210WBS01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044878.D	1		02/10/25 14:29	VX021025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	19.9		0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	20.8		0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	20.7		0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	20.5		0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	42.0		0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	20.6		0.14	0.50	1.00	ug/L
100-42-5	Styrene	21.0		0.16	0.50	1.00	ug/L
75-25-2	Bromoform	20.5		0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	19.8		0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	19.9		0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	19.6		0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	19.9		0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	20.3		0.19	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	46.8		81 - 118		94%	SPK: 50
1868-53-7	Dibromofluoromethane	46.6		80 - 119		93%	SPK: 50
2037-26-5	Toluene-d8	47.0		89 - 112		94%	SPK: 50
460-00-4	4-Bromofluorobenzene	47.2		85 - 114		94%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	144000	5.544				
540-36-3	1,4-Difluorobenzene	262000	6.757				
3114-55-4	Chlorobenzene-d5	225000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	103000	12.018				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	01/28/25	
Project:	CTO WE13		Date Received:	01/30/25	
Client Sample ID:	BP-VPB-192-GW-360-362MS		SDG No.:	Q1250	
Lab Sample ID:	Q1250-07MS		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044889.D	1		02/10/25 18:46	VX021025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	46.2		0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	45.7		0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	47.6		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	52.8		0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	47.0		0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	47.3		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	46.3		0.26	0.75	1.00	ug/L
67-64-1	Acetone	250		1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	45.8		0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	47.7		0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	47.0		0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	47.8		0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	47.6		0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	250		1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	45.0		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	48.0		0.25	0.75	1.00	ug/L
67-66-3	Chloroform	47.3		0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	47.1		0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	45.4		0.19	0.50	1.00	ug/L
71-43-2	Benzene	46.4		0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	47.6		0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	45.6		0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	46.7		0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	47.7		0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	250		0.75	2.50	5.00	ug/L
108-88-3	Toluene	47.4		0.18	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	47.0		0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	47.0		0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	48.6		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	260		1.10	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	01/28/25	
Project:	CTO WE13		Date Received:	01/30/25	
Client Sample ID:	BP-VPB-192-GW-360-362MS		SDG No.:	Q1250	
Lab Sample ID:	Q1250-07MS		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044889.D	1		02/10/25 18:46	VX021025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	47.6		0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	45.5		0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	46.6		0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	47.0		0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	93.5		0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	46.7		0.14	0.50	1.00	ug/L
100-42-5	Styrene	47.9		0.16	0.50	1.00	ug/L
75-25-2	Bromoform	47.4		0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	47.0		0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	47.1		0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	46.0		0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	45.2		0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	47.5		0.19	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	50.8		81 - 118		102%	SPK: 50
1868-53-7	Dibromofluoromethane	49.0		80 - 119		98%	SPK: 50
2037-26-5	Toluene-d8	49.5		89 - 112		99%	SPK: 50
460-00-4	4-Bromofluorobenzene	49.3		85 - 114		99%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	123000	5.55				
540-36-3	1,4-Difluorobenzene	231000	6.757				
3114-55-4	Chlorobenzene-d5	204000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	88500	12.018				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	01/28/25	
Project:	CTO WE13		Date Received:	01/30/25	
Client Sample ID:	BP-VPB-192-GW-360-362MSD		SDG No.:	Q1250	
Lab Sample ID:	Q1250-08MSD		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044890.D	1		02/10/25 19:09	VX021025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	48.9		0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	47.9		0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	50.2		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	52.3		0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	49.5		0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	49.6		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	50.2		0.26	0.75	1.00	ug/L
67-64-1	Acetone	260		1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	49.3		0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	51.1		0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	50.3		0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	50.8		0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	51.2		0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	270		1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	50.5		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	50.7		0.25	0.75	1.00	ug/L
67-66-3	Chloroform	50.8		0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	51.0		0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	49.1		0.19	0.50	1.00	ug/L
71-43-2	Benzene	51.3		0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	51.8		0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	51.2		0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	51.2		0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	52.4		0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	280		0.75	2.50	5.00	ug/L
108-88-3	Toluene	51.8		0.18	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	51.8		0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	52.5		0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	52.6		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	280		1.10	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	01/28/25	
Project:	CTO WE13		Date Received:	01/30/25	
Client Sample ID:	BP-VPB-192-GW-360-362MSD		SDG No.:	Q1250	
Lab Sample ID:	Q1250-08MSD		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044890.D	1		02/10/25 19:09	VX021025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	52.8		0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	48.5		0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	50.5		0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	50.4		0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	100		0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	50.4		0.14	0.50	1.00	ug/L
100-42-5	Styrene	52.6		0.16	0.50	1.00	ug/L
75-25-2	Bromoform	53.2		0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	50.6		0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	50.6		0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	49.9		0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	49.2		0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	51.7		0.19	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	50.2		81 - 118		100%	SPK: 50
1868-53-7	Dibromofluoromethane	50.4		80 - 119		101%	SPK: 50
2037-26-5	Toluene-d8	51.0		89 - 112		102%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.6		85 - 114		103%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	114000	5.544				
540-36-3	1,4-Difluorobenzene	209000	6.757				
3114-55-4	Chlorobenzene-d5	187000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	83000	12.018				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1250 SAS No.: Q1250 SDG No.: Q1250

Instrument ID: MSVOA_X Calibration Date(s): 02/10/2025 02/10/2025

Heated Purge: (Y/N) N Calibration Time(s): 10:25 12:28

GC Column: DB-624UI ID: 0.18 (mm)

LAB FILE ID:								
RRF001 = VX044868.D			RRF005 = VX044869.D			RRF020 = VX044870.D		
RRF050 = VX044871.D			RRF100 = VX044872.D			RRF150 = VX044873.D		
COMPOUND	RRF001	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	% RSD
Chloromethane	0.958	0.860	0.862	0.843	0.805	0.793	0.854	6.9
Vinyl Chloride	0.839	0.846	0.847	0.808	0.807	0.814	0.827	2.3
Bromomethane		0.248	0.252	0.249	0.242	0.246	0.247	1.4
Chloroethane	0.500	0.288	0.280	0.341	0.249	0.181	0.307	35.4
Trichlorofluoromethane	1.062	1.066	1.096	1.029	1.013	1.008	1.046	3.3
1,1,2-Trichlorotrifluoroethane	0.583	0.647	0.668	0.626	0.631	0.637	0.632	4.5
1,1-Dichloroethene	0.647	0.639	0.661	0.630	0.632	0.657	0.644	2
Acetone	0.305	0.292	0.298	0.293	0.285	0.292	0.294	2.3
Carbon Disulfide	1.689	1.732	1.786	1.762	1.789	1.846	1.767	3
Methyl tert-butyl Ether	1.941	2.065	2.130	2.046	2.011	2.110	2.050	3.4
Methylene Chloride	0.747	0.717	0.741	0.704	0.695	0.720	0.721	2.8
trans-1,2-Dichloroethene	0.608	0.622	0.657	0.640	0.633	0.644	0.634	2.7
1,1-Dichloroethane	1.155	1.257	1.292	1.227	1.209	1.257	1.233	3.9
2-Butanone	0.422	0.472	0.504	0.506	0.477	0.487	0.478	6.4
Carbon Tetrachloride	0.457	0.466	0.478	0.453	0.445	0.459	0.460	2.4
cis-1,2-Dichloroethene	0.680	0.783	0.812	0.758	0.758	0.779	0.762	5.9
Chloroform	1.167	1.209	1.268	1.169	1.153	1.208	1.196	3.5
1,1,1-Trichloroethane	1.014	1.003	1.051	1.005	0.984	1.028	1.014	2.3
Methylcyclohexane	0.509	0.571	0.667	0.622	0.634	0.635	0.606	9.4
Benzene	1.370	1.488	1.577	1.470	1.429	1.453	1.465	4.7
1,2-Dichloroethane	0.417	0.465	0.502	0.472	0.462	0.482	0.467	6.1
Trichloroethene	0.293	0.340	0.367	0.335	0.332	0.343	0.335	7.2
1,2-Dichloropropane	0.343	0.354	0.389	0.367	0.360	0.372	0.364	4.3
Bromodichloromethane	0.428	0.481	0.514	0.500	0.500	0.513	0.489	6.6
4-Methyl-2-Pentanone	0.439	0.514	0.562	0.554	0.506	0.498	0.512	8.6
Toluene	0.776	0.872	0.957	0.898	0.866	0.864	0.872	6.7
t-1,3-Dichloropropene	0.417	0.451	0.518	0.514	0.528	0.543	0.495	10
cis-1,3-Dichloropropene	0.452	0.511	0.587	0.577	0.587	0.599	0.552	10.5
1,1,2-Trichloroethane	0.307	0.342	0.362	0.341	0.331	0.329	0.335	5.4
2-Hexanone	0.313	0.360	0.406	0.404	0.369	0.362	0.369	9.3

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1250 SAS No.: Q1250 SDG No.: Q1250

Instrument ID: MSVOA_X Calibration Date(s): 02/10/2025 02/10/2025

Heated Purge: (Y/N) N Calibration Time(s): 10:25 12:28

GC Column: DB-624UI ID: 0.18 (mm)

LAB FILE ID:		RRF001 = VX044868.D		RRF005 = VX044869.D		RRF020 = VX044870.D			
		RRF050 = VX044871.D		RRF100 = VX044872.D		RRF150 = VX044873.D			
COMPOUND	RRF001	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	% RSD	
Dibromochloromethane	0.317	0.342	0.381	0.373	0.368	0.370	0.359	6.7	
Tetrachloroethene	0.306	0.311	0.343	0.310	0.307	0.314	0.315	4.4	
Chlorobenzene	0.969	1.093	1.140	1.096	1.071	1.076	1.074	5.3	
Ethyl Benzene	1.690	1.873	2.021	1.935	1.923	1.929	1.895	5.9	
m/p-Xylenes	0.616	0.700	0.754	0.724	0.706	0.694	0.699	6.6	
o-Xylene	0.661	0.721	0.747	0.707	0.691	0.681	0.701	4.4	
Styrene	0.909	1.124	1.249	1.199	1.161	1.139	1.130	10.4	
Bromoform	0.186	0.247	0.272	0.280	0.276	0.287	0.258	14.7	
Isopropylbenzene	3.735	4.012	4.347	4.045	3.940	4.076	4.026	4.9	
1,1,2,2-Tetrachloroethane	1.429	1.403	1.438	1.366	1.305	1.360	1.383	3.6	
1,3-Dichlorobenzene	1.616	1.669	1.741	1.679	1.663	1.703	1.678	2.5	
1,4-Dichlorobenzene	1.662	1.712	1.762	1.686	1.660	1.701	1.697	2.2	
1,2-Dichlorobenzene	1.512	1.713	1.763	1.666	1.604	1.639	1.650	5.3	
1,2-Dichloroethane-d4		0.764	0.718	0.723	0.707	0.747	0.732	3.2	
Dibromofluoromethane		0.335	0.322	0.320	0.320	0.328	0.325	2	
Toluene-d8		1.239	1.249	1.239	1.208	1.212	1.229	1.5	
4-Bromofluorobenzene		0.404	0.410	0.431	0.415	0.412	0.414	2.5	

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1250 SAS No.: Q1250 SDG No.: Q1250

Instrument ID: MSVOA_X Calibration Date/Time: 02/10/2025 19:32

Lab File ID: VX044891.D Init. Calib. Date(s): 02/10/2025 02/10/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:25 12:28

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Chloromethane	0.854	0.831	0.1	-2.69	50
Vinyl Chloride	0.827	0.785		-5.08	50
Bromomethane	0.247	0.245		-0.81	50
Chloroethane	0.307	0.279		-9.12	50
Trichlorofluoromethane	1.046	1.008		-3.63	50
1,1,2-Trichlorotrifluoroethane	0.632	0.593		-6.17	50
1,1-Dichloroethene	0.644	0.635		-1.4	50
Acetone	0.294	0.310		5.44	50
Carbon Disulfide	1.767	1.726		-2.32	50
Methyl tert-butyl Ether	2.050	2.132		4	50
Methylene Chloride	0.721	0.714		-0.97	50
trans-1,2-Dichloroethene	0.634	0.636		0.31	50
1,1-Dichloroethane	1.233	1.250	0.1	1.38	50
2-Butanone	0.478	0.524		9.62	50
Carbon Tetrachloride	0.460	0.438		-4.78	50
cis-1,2-Dichloroethene	0.762	0.787		3.28	50
Chloroform	1.196	1.210		1.09	50
1,1,1-Trichloroethane	1.014	1.008		-0.59	50
Methylcyclohexane	0.606	0.575		-5.12	50
Benzene	1.465	1.458		-0.48	50
1,2-Dichloroethane	0.467	0.487		4.28	50
Trichloroethene	0.335	0.330		-1.49	50
1,2-Dichloropropane	0.364	0.373		2.47	50
Bromodichloromethane	0.489	0.511		4.5	50
4-Methyl-2-Pentanone	0.512	0.566		10.55	50
Toluene	0.872	0.879		0.8	50
t-1,3-Dichloropropene	0.495	0.512		3.43	50
cis-1,3-Dichloropropene	0.552	0.577		4.53	50
1,1,2-Trichloroethane	0.335	0.350		4.48	50
2-Hexanone	0.369	0.413		11.92	50
Dibromochloromethane	0.359	0.384		6.96	50
Tetrachloroethene	0.315	0.293		-6.98	50
Chlorobenzene	1.074	1.078	0.3	0.37	50
Ethyl Benzene	1.895	1.898		0.16	50
m/p-Xylenes	0.699	0.704		0.71	50
o-Xylene	0.701	0.707		0.86	50
Styrene	1.130	1.169		3.45	50
Bromoform	0.258	0.270	0.1	4.65	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1250 SAS No.: Q1250 SDG No.: Q1250

Instrument ID: MSVOA_X Calibration Date/Time: 02/10/2025 19:32

Lab File ID: VX044891.D Init. Calib. Date(s): 02/10/2025 02/10/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:25 12:28

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Isopropylbenzene	4.026	4.020		-0.15	50
1,1,2,2-Tetrachloroethane	1.383	1.415	0.3	2.31	50
1,3-Dichlorobenzene	1.678	1.664		-0.83	50
1,4-Dichlorobenzene	1.697	1.666		-1.83	50
1,2-Dichlorobenzene	1.650	1.683		2	50
1,2-Dichloroethane-d4	0.732	0.679		-7.24	50
Dibromofluoromethane	0.325	0.294		-9.54	50
Toluene-d8	1.229	1.099		-10.58	50
4-Bromofluorobenzene	0.414	0.382		-7.73	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.