

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q1250 SAS No.: Q1250 SDG NO.: Q1250
Lab File ID: VX044871.D Date Analyzed: 02/10/2025
Instrument ID: MSVOA_X Time Analyzed: 11:34
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	127984	5.54	233736	6.76	204751	10.05
UPPER LIMIT	255968	6.043	467472	7.257	409502	10.549
LOWER LIMIT	63992	5.043	116868	6.257	102376	9.549
EPA SAMPLE NO.						
BP-VPB-192-TB-20250127	92731	5.54	189692	6.76	170630	10.05
BP-VPB-192-GW-420-422	96890	5.55	196670	6.76	180490	10.05
BP-VPB-192-GW-300-302	99257	5.55	200954	6.76	183816	10.05
BP-VPB-192-GW-320-322	96132	5.55	194678	6.76	175107	10.05
BP-VPB-192-GW-340-342	96963	5.54	193723	6.76	175745	10.05
BP-VPB-192-GW-360-362	93911	5.54	187948	6.76	173877	10.05
BP-VPB-192-GW-360-362MS	123128	5.55	231125	6.76	204212	10.05
BP-VPB-192-GW-360-362MSD	113791	5.54	209107	6.76	187126	10.05
BP-VPB-192-GW-380-382	98000	5.54	197272	6.76	182089	10.05
BP-VPB-192-GW-400-402	94362	5.55	191816	6.76	178953	10.05
BP-VPB-192-GW-440-442	91120	5.54	183168	6.76	166878	10.05
VX0210WBL01	95840	5.54	200206	6.76	182254	10.05
VX0210WBS01	143675	5.54	262165	6.76	224717	10.05

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.