

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
 Lab Code: CHEM Case No.: Q1250 SAS No.: Q1250 SDG NO.: Q1250
 Lab File ID: VX044871.D Date Analyzed: 02/10/2025
 Instrument ID: MSVOA_X Time Analyzed: 11:34
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	92648	12.018				
UPPER LIMIT	185296	12.518				
LOWER LIMIT	46324	11.518				
EPA SAMPLE NO.						
BP-VPB-192-TB-20250127	69234	12.02				
BP-VPB-192-GW-420-422	78413	12.02				
BP-VPB-192-GW-300-302	78851	12.02				
BP-VPB-192-GW-320-322	76744	12.02				
BP-VPB-192-GW-340-342	75903	12.02				
BP-VPB-192-GW-360-362	73228	12.02				
BP-VPB-192-GW-360-362MS	88501	12.02				
BP-VPB-192-GW-360-362MSD	83002	12.02				
BP-VPB-192-GW-380-382	77317	12.02				
BP-VPB-192-GW-400-402	76716	12.02				
BP-VPB-192-GW-440-442	68152	12.02				
VX0210WBL01	76715	12.02				
VX0210WBS01	103194	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.