

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB134591

Review By	Kareem Khairalla	Review On	2/7/2025 3:44:51 PM
Supervise By	Janvi Patel	Supervise On	2/10/2025 1:30:43 PM
STD. NAME	STD REF.#		
ICAL Standard	MP84204,MP84381,MP84383,MP84384,MP84385,MP84386,MP84388		
ICV Standard	MP84389		
CCV Standard	MP84392		
ICSA Standard	MP84390,MP84391		
CRI Standard			
LCS Standard			
Chk Standard	MP84218,MP84219		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	02/05/25 15:17		Kareem	OK
2	S1	S01	CAL2	02/05/25 15:21		Kareem	OK
3	S2	S02	CAL3	02/05/25 15:25		Kareem	OK
4	S3	S03	CAL4	02/05/25 15:29		Kareem	OK
5	S4	S04	CAL5	02/05/25 15:33		Kareem	OK
6	S5	S05	CAL6	02/05/25 15:37		Kareem	OK
7	S6	S06	CAL7	02/05/25 15:41		Kareem	OK
8	ICV001	ICV001	ICV	02/05/25 17:08	Inst. paused for calibration review	Kareem	OK
9	ICB001	ICB001	ICB	02/05/25 17:15	Inst. paused for review	Kareem	OK
10	ICSA001	ICSA001	ICSA	02/05/25 17:41	Inst. paused for review	Kareem	OK
11	ICSAB001	ICSAB001	ICSAB	02/05/25 17:46		Kareem	OK
12	CCV001	CCV001	CCV	02/05/25 18:10	Inst. paused for review	Kareem	OK
13	CCB001	CCB001	CCB	02/05/25 18:14		Kareem	OK
14	PB166534BL	PBS534	MB	02/05/25 18:18		Kareem	OK
15	PB166534BS	LCS534	LCS	02/05/25 18:22		Kareem	OK
16	PB166531BL	PBS531	MB	02/05/25 18:26		Kareem	OK
17	Q1251-01	E1491	SAM	02/05/25 18:34		Kareem	OK
18	Q1251-02	E1492	SAM	02/05/25 18:38		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB134591

Review By	Kareem Khairalla	Review On	2/7/2025 3:44:51 PM
Supervise By	Janvi Patel	Supervise On	2/10/2025 1:30:43 PM
STD. NAME	STD REF.#		
ICAL Standard	MP84204,MP84381,MP84383,MP84384,MP84385,MP84386,MP84388		
ICV Standard	MP84389		
CCV Standard	MP84392		
ICSA Standard	MP84390,MP84391		
CRI Standard			
LCS Standard			
Chk Standard	MP84218,MP84219		

19	Q1251-03	E1493	SAM	02/05/25 18:41		Kareem	OK
20	PB166531BS	LCS531	LCS	02/05/25 18:45		Kareem	OK
21	Q1251-04	E1494	SAM	02/05/25 18:49		Kareem	OK
22	Q1251-05	E1495	SAM	02/05/25 18:53		Kareem	OK
23	Q1251-06	E1496	SAM	02/05/25 18:57		Kareem	OK
24	Q1251-07	E1496D	DUP	02/05/25 19:00		Kareem	OK
25	Q1251-06L	E1496L	SD	02/05/25 19:04		Kareem	OK
26	Q1251-08	E1496S	MS	02/05/25 19:08		Kareem	OK
27	Q1251-09	E1497	SAM	02/05/25 19:12		Kareem	OK
28	Q1272-01	MBHHA0	SAM	02/05/25 19:16		Kareem	OK
29	Q1272-02	MBHHA1	SAM	02/05/25 19:19		Kareem	OK
30	Q1272-03	MBHHA2	SAM	02/05/25 19:23		Kareem	OK
31	Q1272-04	MBHHA2D	DUP	02/05/25 19:27		Kareem	OK
32	Q1272-03L	MBHHA2L	SD	02/05/25 19:31		Kareem	OK
33	Q1272-05	MBHHA2S	MS	02/05/25 19:35	MS fail for Zn (898.23 ppb), Ag & Sb,Tl(Below RL)	Kareem	OK
34	CCV002	CCV002	CCV	02/05/25 19:54		Kareem	OK
35	CCB002	CCB002	CCB	02/05/25 19:58		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB134617

Review By	Kareem Khairalla	Review On	2/7/2025 3:55:23 PM
Supervise By	Janvi Patel	Supervise On	2/10/2025 1:38:43 PM
STD. NAME	STD REF.#		
ICAL Standard	MP84204,MP84388,MP84386,MP84385,MP84384,MP84383,MP84381		
ICV Standard	MP84389		
CCV Standard	MP84392		
ICSA Standard	MP84390,MP84391		
CRI Standard			
LCS Standard			
Chk Standard	MP84218,MP84219		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	02/06/25 12:47		Kareem	OK
2	S1	S01	CAL2	02/06/25 12:51		Kareem	OK
3	S2	S02	CAL3	02/06/25 12:55		Kareem	OK
4	S3	S03	CAL4	02/06/25 12:59		Kareem	OK
5	S4	S04	CAL5	02/06/25 13:03		Kareem	OK
6	S5	S05	CAL6	02/06/25 13:07		Kareem	OK
7	S6	S06	CAL7	02/06/25 13:11		Kareem	OK
8	ICV002	ICV002	ICV	02/06/25 14:09	Inst. paused for calibration review	Kareem	OK
9	ICB002	ICB002	ICB	02/06/25 14:14		Kareem	OK
10	ICSA002	ICSA002	ICSA	02/06/25 14:18		Kareem	OK
11	ICSAB002	ICSAB002	ICSAB	02/06/25 14:25	Inst. paused for review	Kareem	OK
12	CCV003	CCV003	CCV	02/06/25 14:29		Kareem	OK
13	CCB003	CCB003	CCB	02/06/25 14:32		Kareem	OK
14	Q1151-13	MC0B02	SAM	02/06/25 14:36		Kareem	OK
15	Q1251-10	E1498	SAM	02/06/25 14:41		Kareem	OK
16	PB166571BL	PBW571	MB	02/06/25 14:45		Kareem	OK
17	PB166571BS	LCS571	LCS	02/06/25 14:49		Kareem	OK
18	PB166568BL	PBW568	MB	02/06/25 14:53		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB134617

Review By	Kareem Khairalla	Review On	2/7/2025 3:55:23 PM
Supervise By	Janvi Patel	Supervise On	2/10/2025 1:38:43 PM

STD. NAME	STD REF.#
ICAL Standard	MP84204,MP84388,MP84386,MP84385,MP84384,MP84383,MP84381
ICV Standard	MP84389
CCV Standard	MP84392
ICSA Standard	MP84390,MP84391
CRI Standard	
LCS Standard	
Chk Standard	MP84218,MP84219

19	PB166568BS	LCS568	LCS	02/06/25 14:57		Kareem	OK
20	PB166494TB	LEB494	MB	02/06/25 15:01		Kareem	OK
21	Q1272-03A	MBHHA2A	PS	02/06/25 15:05	NOT USE	Kareem	Not Ok
22	Q1273-01	MBHHA0	SAM	02/06/25 15:08		Kareem	OK
23	Q1273-02	MBHHA1	SAM	02/06/25 15:12		Kareem	OK
24	Q1273-03	MBHHA2	SAM	02/06/25 15:16	NOT USE	Kareem	Not Ok
25	Q1273-04	MBHHA2D	DUP	02/06/25 15:20	NOT USE	Kareem	Not Ok
26	Q1273-03L	MBHHA2L	SD	02/06/25 15:24	NOT USE	Kareem	Not Ok
27	Q1273-05	MBHHA2S	MS	02/06/25 15:28	NOT USE	Kareem	Not Ok
28	CCV004	CCV004	CCV	02/06/25 15:31		Kareem	OK
29	CCB004	CCB004	CCB	02/06/25 15:35		Kareem	OK