

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51981 MA No.: SDG No.: E1491
Instrument ID: P5 Start Date: 02/05/2025
Analytical Method: ICP-AES Run Batch: LB134591
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Arsenic	0	-0	0	10	11	10	1250	1300	4
Lead	0	0.0	0	10	8.7	-13	6250	6100	-2

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51981 MA No.: SDG No.: E1491
Instrument ID: P5 Start Date: 02/05/2025
Analytical Method: ICP-AES Run Batch: LB134591
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Arsenic	2500	2600	4	5000	5100	2	10000	9700	-3
Lead	12500	13000	4	25000	25000	0	50000	50000	0

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51981 MA No.: _____ SDG No.: E1491
Instrument ID: P5 Start Date: 02/06/2025
Analytical Method: ICP-AES Run Batch: LB134617
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Arsenic	0	-0	0	10	12	20	1250	1300	4
Lead	0	-0	0	10	10	0	6250	6100	-2

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51981 MA No.: SDG No.: E1491
Instrument ID: P5 Start Date: 02/06/2025
Analytical Method: ICP-AES Run Batch: LB134617
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Arsenic	2500	2600	4	5000	5100	2	10000	9800	-2
Lead	12500	13000	4	25000	25000	0	50000	49000	-2