

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011

Lab Code: ACE Case No.: 51981 MA No. : SDG No.: E1491

Initial Calibration Verification Source : MP84389

Continuing Calibration Verification Source : MP84392

Run Batch: LB134591 Analytical Method: ICP-AES

Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV001				ID: CCV001				ID: CCV002		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Arsenic	1000	1000	104	0	5000	4700	94	0	4800	96	0
Lead	1000	950	95	0	25000	24000	97	0	25000	100	0

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011

Lab Code: ACE Case No.: 51981 MA No. : SDG No.: E1491

Initial Calibration Verification Source : MP84389

Continuing Calibration Verification Source : MP84392

Run Batch: LB134617 Analytical Method: ICP-AES

Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV002				ID: CCV003				ID: CCV004		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Arsenic	1000	1100	108	0	5000	4900	98	0	4900	99	0
Lead	1000	990	99	0	25000	25000	101	0	25000	101	0