

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: Q1252

Project: Pre Treatment Plant

RunNo.: LB134510

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Orthophosphate as P	mg/L	0.509	0.50	102	90-110	01/31/2025
Sample ID: CCV1 Orthophosphate as P	mg/L	0.512	0.5	102	90-110	01/31/2025
Sample ID: CCV2 Orthophosphate as P	mg/L	0.509	0.5	102	90-110	01/31/2025

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: Q1252

Project: Pre Treatment Plant

RunNo.: LB134512

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Phosphorus, Total	mg/L	0.504	0.50	101	90-110	01/31/2025
Sample ID: CCV1 Phosphorus, Total	mg/L	0.509	0.50	102	90-110	01/31/2025
Sample ID: CCV2 Phosphorus, Total	mg/L	0.504	0.50	101	90-110	01/31/2025

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: Q1252

Project: Pre Treatment Plant

RunNo.: LB134532

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	0.99	1	99	90-110	02/03/2025
Sample ID: CCV1 Ammonia as N	mg/L	0.94	1	94	90-110	02/03/2025
Sample ID: CCV2 Ammonia as N	mg/L	0.97	1	97	90-110	02/03/2025
Sample ID: CCV3 Ammonia as N	mg/L	0.95	1	95	90-110	02/03/2025
Sample ID: CCV4 Ammonia as N	mg/L	0.99	1	99	90-110	02/03/2025

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: Q1252

Project: Pre Treatment Plant

RunNo.: LB134532

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------