

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q1252

Project: Pre Treatment Plant

Run No.: LB134509

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134509BS							
BOD5	mg/L	198	197		100	1	84.6-115.4	01/31/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q1252

Project: Pre Treatment Plant

Run No.: LB134510

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134510BS							
Orthophosphate as P	mg/L	0.5	0.49		99	1	90-110	01/31/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q1252

Project: Pre Treatment Plant

Run No.: LB134521

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134521BS							
Oil and Grease	mg/L	20.0	16.9		84	1	78-114	02/01/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q1252

Project: Pre Treatment Plant

Run No.: LB134570

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134570BS							
TSS	mg/L	550	532		97	1	90-110	02/05/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q1252

Project: Pre Treatment Plant

Run No.: LB134512

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB166431BS							
Phosphorus, Total	mg/L	0.50	0.51		101	1	90-110	01/31/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q1252

Project: Pre Treatment Plant

Run No.: LB134532

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB166477BS							
Ammonia as N	mg/L	1	0.97		97	1	90-110	02/03/2025