

FORM 6 - IN
DUPLICATES

Lab Name: Chemtech Consulting Group Contract: EPW14030
 Lab Code: ACE Case No.: 51967 MA No. : SDG No.: MDCZM3
 Matrix : Water Analytical Method: ICP-MS
 % Solids:
 Concentration Units (µg/L, mg/L or mg/kg dry weight): ug/L

Analyte	Control Limit	Sample (S)	Q	Duplicate (D)	Q	RPD	Q
Aluminum	20.0	33		32		2	
Antimony		2.0	U	2.0	U		
Arsenic		0.35	J	0.31	J		
Barium		87		86		1	
Beryllium		0.28	J	0.29	J		
Cadmium		1.0	U	1.0	U		
Calcium		5500		5500		1	
Chromium		0.64	J	1.0	J		
Cobalt	1.0	3.5		3.5		1	
Copper		2.0	U	2.0	U		
Iron	200	210		210		3	
Lead		1.0	U	1.0	U		
Magnesium	500	1700		1600		2	
Manganese		210		210		3	
Nickel		9.0		9.0		0	
Potassium	500	790		790		0	
Selenium		5.0	U	5.0	U		
Silver		1.0	U	1.0	U		
Sodium		28000		28000		1	
Thallium		1.0	U	1.0	U		
Vanadium		0.21	J	0.22	J		
Zinc	5.0	12		12		0	