



**284 Sheffield Street  
Mountainside, NJ 07092**

## **SDG NARRATIVE**

**USEPA**

**SDG # MDCZM3**

**CASE # 51967**

**CONTRACT # 68HERH20D0011**

**SOW# SFAM01.1**

**LAB NAME: Alliance Technical Group, LLC**

**LAB CODE: ACE**

**LAB ORDER ID # Q1255**

### **A. Number of Samples and Date of Receipt**

14 Water samples were delivered to the laboratory intact on 01/31/2025, 02/01/2025.

### **B. Parameters**

Test requested for Metals CLP MS = Aluminum, Antimony, Arsenic, Barium, Beryllium, Cadmium, Calcium, Chromium, Cobalt, Copper, Iron, Lead, Magnesium, Manganese, Nickel, Potassium, Selenium, Silver, Sodium, Thallium, Vanadium, Zinc, Mercury.

### **C. Cooler Temp**

Indicator Bottle: Presence/Absence

Cooler: 2.1°C, 2.0°C, 1.8°C, 1.5°C.

### **D. Detail Documentation (related to Sample Handling Shipping, Analytical Problem, Temp of Cooler etc):**

Issue 1: A "P" or "M" prefix was listed at the beginning of a CLP sample ID.

### **E. Corrective Action taken for above:**

Resolution 1: To maintain COC integrity, ASB requests no changes to the Sample IDs. The laboratory will note the issue in the SDG Narrative and proceed with the analysis of the samples.

### **F. Analytical Techniques:**

All analyses were based on CLP Methodology by method SFAM01.1.



**284 Sheffield Street  
Mountainside, NJ 07092**

**G. Calculation:**

**Calculation for ICP-MS Water Sample:**

$$\text{Concentration or Result } (\mu\text{g/L}) = C \times \frac{V_f}{V_i} \times \text{DF}$$

Where,

C = Instrument value in ppb (The average of all replicate integrations)

V<sub>f</sub> = Final digestion volume (mL)

V<sub>i</sub> = Initial aliquot amount (mL) (Sample amount taken in prep)

DF = Dilution Factor

**Example Calculation For Sample MDCZM3 For Aluminum:**

If C = 11.25 ppb

V<sub>f</sub> = 50 ml

V<sub>i</sub> = 50 ml

DF = 1

$$\text{Concentration or Result } (\mu\text{g/L}) = 11.25 \times \frac{50}{50} \times 1$$

$$= 11.25 \mu\text{g/L}$$

$$= 11 \mu\text{g/L} \text{ (Reported Result with Signification)}$$

**Calculation for Hg Water Sample:**

$$\text{Concentration or Result } (\mu\text{g/L}) = C \times \text{DF}$$

Where,

C = Instrument response in  $\mu\text{g/L}$  from the calibration curve.

DF = Dilution Factor

**Example Calculation for MDCZP0:**

If C = 0.0381 ppb

DF = 1

$$\text{Concentration or Result } (\mu\text{g/L}) = 0.0381 \times 1$$

$$= 0.0381 \mu\text{g/L}$$

$$= 0.038 \mu\text{g/L} \text{ (Reported Result with Signification)}$$



**284 Sheffield Street  
Mountainside, NJ 07092**

#### **H. QA/ QC**

Calibrations met requirements. Interference check met requirements. Blank analyses did not indicate any presence of contamination. Laboratory Control sample was within control limits. Spike sample did meet requirements except for Selenium. Duplicate sample did meet requirements. Serial Dilution did meet requirements.

Collision cell is being used to remove potential interferences. The analytes Na, Mg, Al, K, V, Cr, Mn, Fe, Co, Ni, Cu, Zn, As are being analyzed with collision cell and analytes Be, B, Ca, Ti, Se, Sr, Zr, Mo, Ag, Cd, Sn, Sb, Ba, Tl, Pb, U are being analyzed with Non-Collision Cell. Helium gas is used for the Collision Cell analysis.

Internal Standard Association for ICP-MS analysis.

Target Analyte	Associated Internal Standard
Aluminum	45Sc
Antimony	159Tb
Arsenic	89Y
Barium	159Tb
Beryllium	6Li
Cadmium	159Tb
Calcium	45Sc
Chromium	45Sc
Cobalt	45Sc
Copper	45Sc
Iron	45Sc
Lead	209Bi
Magnesium	45Sc
Manganese	45Sc
Nickel	45Sc
Potassium	45Sc



**284 Sheffield Street  
Mountainside, NJ 07092**

Selenium	89Y
Silver	159Tb
Sodium	45Sc
Thallium	209Bi
Vanadium	45Sc
Zinc	45Sc

I certify that the data package is in compliance with the terms and conditions of the contract both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the Laboratory Director or his designee, as verified by the following signature.

Signature\_\_\_\_\_

Name: Nimisha Pandya

Date\_\_\_\_\_

Title: Document Control Officer