

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: GFEL01

Lab Code: CHEM Case No.: Q1256 SAS No.: Q1256 SDG No.: Q1256

Instrument ID: MSVOA_L Calibration Date/Time: 02/03/2025 10:48

Lab File ID: VL041958.D Init. Calib. Date(s): 01/14/2025 01/14/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 12:55 16:05

GC Column: RTX-1 ID: 0.32 (mm)

COMPOUND	RRF	RRF010	MIN RRF	%D	MAX%D
Vinyl Chloride	0.519	0.529		1.93	30
1,1-Dichloroethene	0.501	0.473		-5.59	30
cis-1,2-Dichloroethene	1.349	1.325		-1.78	30
1,1,1-Trichloroethane	1.938	2.024		4.44	30
Trichloroethene	0.364	0.408		12.09	30
Tetrachloroethene	0.318	0.325		2.2	30
1-Bromo-4-Fluorobenzene	0.788	0.780		-1.01	30

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.