

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL020325\
 Data File : VL041958.D
 Acq On : 03 Feb 2025 10:48
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Feb 04 02:30:31 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL011425AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Tue Jan 14 23:15:42 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	103	0.00
2 T	Dichlorodifluoromethane	1.362	1.387	-1.8	110	0.00
3	Chlorodifluoromethane	1.427	1.469	-2.9	114	0.00
4	Chloromethane	0.512	0.549	-7.2	115	0.00
5 T	Vinyl Chloride	0.519	0.529	-1.9	117	0.00
6 T	Bromomethane	0.247	0.243	1.6	110	0.00
7	Chloroethane	0.202	0.217	-7.4	116	0.00
8 T	Dichlorotetrafluoroethane	1.190	1.183	0.6	112	0.00
9 T	Propene	0.588	0.609	-3.6	111	0.00
10 T	Heptane	1.571	1.695	-7.9	115	-0.02
11 T	Trichlorofluoromethane	1.417	1.348	4.9	104	0.00
12 T	1,1,2-Trichlorotrifluoroeth	1.110	1.035	6.8	101	0.00
13	Ethanol	0.043	0.028#	34.9#	106	0.00
14 T	Bromoethene	0.407	0.385	5.4	107	0.00
15 T	Acetone	1.201	1.121	6.7	117	0.00
16 T	1,3-Butadiene	0.539	0.551	-2.2	115	0.00
17	tert-Butyl alcohol	1.478	1.312	11.2	97	0.00
18 T	1,1-Dichloroethene	0.501	0.473	5.6	107	0.00
19 T	Isopropyl Alcohol	0.712	0.641	10.0	107	0.00
20 T	Methylene Chloride	0.475	0.400	15.8	106	0.00
21 T	Allyl Chloride	0.832	0.805	3.2	105	0.00
22 T	trans-1,2-Dichloroethene	0.565	0.548	3.0	105	0.00
23 T	Vinyl Acetate	0.595	0.536	9.9	103	0.00
24 T	1,1-Dichloroethane	1.093	1.089	0.4	111	0.00
25 T	Ethyl Acetate	3.239	3.507	-8.3	117	0.00
26 T	Hexane	1.281	1.372	-7.1	113	0.00
27 T	Carbon Disulfide	1.268	1.273	-0.4	105	0.00
28 T	Methyl tert-Butyl Ether	0.979	0.952	2.8	127	0.00
29 T	Chloroform	2.025	2.120	-4.7	115	0.00
30 T	Cyclohexane	1.111	1.163	-4.7	111	0.00
31 T	cis-1,2-Dichloroethene	1.349	1.325	1.8	108	0.00
32 T	1,1,1-Trichloroethane	1.938	2.024	-4.4	110	0.00
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	96	-0.01
34 T	2-Butanone	0.636	0.729	-14.6	116	0.00
35 T	Carbon Tetrachloride	0.628	0.747	-18.9	110	-0.01
36 T	Benzene	0.917	1.035	-12.9	112	0.00
37 T	1,2-Dichloroethane	0.515	0.562	-9.1	109	0.00
38 T	Trichloroethene	0.364	0.408	-12.1	110	-0.01
39 T	1,2-Dichloropropane	0.322	0.362	-12.4	113	0.00
40 T	1,4-Dioxane	0.162	0.174	-7.4	113	-0.02
41 T	Tetrahydrofuran	0.346	0.406	-17.3	112	0.00
42 T	Bromodichloromethane	0.691	0.803	-16.2	112	-0.01
43	Methyl Methacrylate	0.341	0.364	-6.7	103	0.00
44 T	2,2,4-Trimethylpentane	1.484	1.749	-17.9	115	-0.01
45 T	t-1,3-Dichloropropene	0.319	0.345	-8.2	97	0.00
46 T	cis-1,3-Dichloropropene	0.419	0.472	-12.6	101	-0.01
47 T	1,1,2-Trichloroethane	0.341	0.389	-14.1	112	0.00
48 T	Dibromochloromethane	0.526	0.627	-19.2	110	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL020325\
 Data File : VL041958.D
 Acq On : 03 Feb 2025 10:48
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Feb 04 02:30:31 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL011425AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Tue Jan 14 23:15:42 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	Bromoform	0.432	0.524	-21.3	109	0.00
50 T	4-Methyl-2-Pentanone	0.868	0.998	-15.0	115	-0.02
51 T	2-Hexanone	0.667	0.784	-17.5	115	0.00
52 T	Tetrachloroethene	0.318	0.325	-2.2	106	0.00
53 T	Toluene	0.999	1.126	-12.7	107	-0.01
54 T	1,2-Dibromoethane	0.469	0.534	-13.9	112	0.00
55 I	Chlorobenzene-d5	1.000	1.000	0.0	93	0.00
56	1,1,1,2-Tetrachloroethane	0.499	0.593	-18.8	114	0.00
57 T	Chlorobenzene	0.922	1.041	-12.9	114	0.00
58 T	Ethyl Benzene	1.563	1.802	-15.3	108	0.00
59 T	m/p-Xylene	1.253	1.490	-18.9	113	0.00
60 T	o-Xylene	1.243	1.495	-20.3	113	0.00
61 T	Styrene	0.558	0.663	-18.8	105	0.00
62	Isopropylbenzene	1.924	2.247	-16.8	112	0.00
63 T	1,1,2,2-Tetrachloroethane	0.795	0.976	-22.8	121	0.00
64	n-propylbenzene	0.486	0.581	-19.5	115	0.00
65	tert-Butylbenzene	1.753	2.010	-14.7	111	0.00
66 T	Benzyl Chloride	0.109	0.121	-11.0	98	0.00
67	sec-Butylbenzene	2.414	2.814	-16.6	114	0.00
68 S	1-Bromo-4-Fluorobenzene	0.788	0.780	1.0	92	0.00
69	p-Isopropyltoluene	1.988	2.337	-17.6	113	0.00
70	n-Butylbenzene	1.972	2.285	-15.9	112	0.00
71	2-Chlorotoluene	1.423	1.657	-16.4	111	0.00
72 T	4-Ethyltoluene	1.487	1.774	-19.3	112	0.00
73 T	1,3,5-Trimethylbenzene	1.313	1.525	-16.1	112	0.00
74 T	1,2,4-Trimethylbenzene	1.422	1.659	-16.7	114	0.00
75 T	1,3-Dichlorobenzene	0.843	0.964	-14.4	115	0.00
76 T	1,4-Dichlorobenzene	0.825	0.948	-14.9	114	0.00
77 T	1,2-Dichlorobenzene	0.835	0.957	-14.6	117	0.00
78 T	Hexachloro-1,3-Butadiene	0.686	0.667	2.8	106	0.00
79 T	Naphthalene	1.432	1.490	-4.1	110	0.00
80 T	Naphthalene,2-methyl-	0.679	0.468	31.1#	94	0.00
81 T	1,2,4-Trichlorobenzene	0.699	0.676	3.3	105	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0