



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_L\Data\VL020325\  
Data File : VL041957.D  
Acq On : 03 Feb 2025 09:47  
Operator : SY/MD  
Sample : BFB  
Misc : 400mL/MSVOA\_L  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_L  
ClientSampleId :  
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA\_L\methods\VL011425AIR.M  
Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA\_L Fri Aug 26 06:05:16 2022  
Last Update : Tue Jan 14 23:15:42 2025



AutoFind: Scans 3178, 3179, 3180; Background Corrected with Scan 3165

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 50             | 95              | 8               | 40              | 22.6         | 51539      | PASS                |
| 75             | 95              | 30              | 66              | 54.7         | 125019     | PASS                |
| 95             | 95              | 100             | 100             | 100.0        | 228352     | PASS                |
| 96             | 95              | 5               | 9               | 6.8          | 15439      | PASS                |
| 173            | 174             | 0.00            | 2               | 0.0          | 0          | PASS                |
| 174            | 95              | 50              | 120             | 58.4         | 133352     | PASS                |
| 175            | 174             | 4               | 9               | 7.7          | 10244      | PASS                |
| 176            | 174             | 93              | 101             | 98.0         | 130661     | PASS                |
| 177            | 176             | 5               | 9               | 6.6          | 8641       | PASS                |

Average of 10.374 to 10.380 min.: VL041957.D\data.ms

BFB

Modified:subtracted

| m/z   | abund. | m/z   | abund. | m/z   | abund. | m/z   | abund. |
|-------|--------|-------|--------|-------|--------|-------|--------|
| 33.10 | 1120   | 45.05 | 2463   | 55.00 | 689    | 65.10 | 841    |
| 36.05 | 2514   | 46.00 | 184    | 56.05 | 3741   | 66.05 | 6      |
| 37.05 | 11892  | 47.00 | 2084   | 57.05 | 7026   | 67.10 | 72     |
| 38.05 | 11097  | 47.90 | 958    | 58.05 | 254    | 68.10 | 2382   |
| 39.10 | 4832   | 48.10 | 422    | 58.90 | 19     | 69.10 | 24107  |
| 39.95 | 251    | 49.05 | 10474  | 60.05 | 2002   | 70.05 | 1798   |
| 40.90 | 35     | 50.10 | 51539  | 61.05 | 11945  | 71.10 | 45     |
| 42.20 | 26     | 51.05 | 15057  | 62.05 | 11539  | 72.05 | 1431   |
| 42.80 | 21     | 51.90 | 324    | 63.10 | 8334   | 73.05 | 10025  |
| 43.15 | 208    | 52.10 | 274    | 63.90 | 422    | 74.10 | 39931  |
| 44.05 | 1487   | 52.90 | 142    | 64.15 | 736    | 75.10 | 125019 |

Average of 10.374 to 10.380 min.: VL041957.D\data.ms

BFB

Modified:subtracted

| m/z   | abund. | m/z   | abund. | m/z    | abund. | m/z    | abund. |
|-------|--------|-------|--------|--------|--------|--------|--------|
| 76.10 | 10646  | 88.00 | 7333   | 99.00  | 28     | 111.80 | 156    |
| 77.05 | 1418   | 88.70 | 126    | 100.80 | 23     | 112.10 | 61     |
| 77.80 | 192    | 89.60 | 71     | 103.05 | 125    | 112.95 | 283    |
| 77.95 | 720    | 90.80 | 252    | 104.05 | 947    | 114.75 | 115    |
| 78.95 | 5171   | 91.05 | 573    | 105.00 | 301    | 115.10 | 34     |
| 80.00 | 1716   | 92.10 | 6387   | 105.85 | 897    | 115.90 | 824    |
| 80.95 | 5529   | 93.05 | 8721   | 106.85 | 269    | 116.90 | 1318   |
| 82.05 | 1302   | 94.10 | 26645  | 107.10 | 82     | 117.95 | 645    |
| 82.95 | 83     | 95.10 | 228352 | 109.75 | 66     | 118.95 | 803    |
| 86.05 | 386    | 96.10 | 15439  | 110.35 | 48     | 119.90 | 29     |
| 87.05 | 8233   | 97.05 | 363    | 110.85 | 100    | 121.70 | 49     |

Average of 10.374 to 10.380 min.: VL041957.D\data.ms

BFB

Modified:subtracted

| m/z    | abund. | m/z    | abund. | m/z    | abund. | m/z    | abund. |
|--------|--------|--------|--------|--------|--------|--------|--------|
| 123.85 | 123    | 136.75 | 135    | 148.95 | 203    | 160.85 | 491    |
| 124.80 | 43     | 137.00 | 198    | 149.90 | 229    | 165.00 | 18     |
| 125.70 | 49     | 140.05 | 224    | 151.65 | 78     | 168.80 | 18     |
| 127.95 | 809    | 141.00 | 1633   | 153.15 | 149    | 169.95 | 60     |
| 128.90 | 210    | 141.90 | 327    | 154.05 | 160    | 170.60 | 29     |
| 129.10 | 123    | 142.95 | 2114   | 154.95 | 531    | 170.90 | 161    |
| 129.90 | 218    | 144.00 | 72     | 155.60 | 22     | 171.85 | 664    |
| 130.05 | 579    | 145.00 | 241    | 156.90 | 253    | 174.05 | 133352 |
| 130.70 | 97     | 145.90 | 188    | 157.10 | 161    | 175.10 | 10244  |
| 131.00 | 189    | 146.95 | 190    | 157.80 | 111    | 176.05 | 130661 |
| 134.90 | 481    | 148.00 | 461    | 158.85 | 171    | 177.05 | 8641   |

Average of 10.374 to 10.380 min.: VL041957.D\data.ms

BFB

Modified:subtracted

| m/z    | abund. | m/z | abund. | m/z | abund. | m/z | abund. |
|--------|--------|-----|--------|-----|--------|-----|--------|
| 178.00 | 253    |     |        |     |        |     |        |
| 263.90 | 65     |     |        |     |        |     |        |

Instrument :

MSVOA\_L

ClientSampleId :

BFB