

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL020325\
 Data File : VL041961.D
 Acq On : 03 Feb 2025 12:57
 Operator : SY/MD
 Sample : VL0203ABS01
 Misc : 400mL/MSVOA_L
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0203ABS01

Quant Time: Feb 04 02:34:05 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL011425AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Tue Jan 14 23:15:42 2025

Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 02/04/2025
 Supervised By : Mahesh Dadoda 02/04/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.790	49	152847	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.965	114	418895	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.882	117	367437	10.000	ppbv	-0.01

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	10.380	95	287209	9.924	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.200%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.499	85	203280	9.764	ppbv	99
3) Chlorodifluoromethane	1.476	51	225433	10.336	ppbv	99
4) Chloromethane	1.535	50	81880	10.457	ppbv	94
5) Vinyl Chloride	1.583	62	77713	9.799	ppbv	98
6) Bromomethane	1.670	94	36640	9.694	ppbv	88
7) Chloroethane	1.706	64	31280	10.137	ppbv	94
8) Dichlorotetrafluoroethane	1.557	85	179670	9.875	ppbv	98
9) Propene	1.486	41	91411	10.174	ppbv	99
10) Heptane	4.985	43	161638	6.731	ppbv	99
11) Trichlorofluoromethane	1.881	101	201683	9.312	ppbv	96
12) 1,1,2-Trichlorotrifluo...	2.143	101	164938	9.717	ppbv	98
13) Ethanol	1.722	45	4592	6.978	ppbv	91
14) Bromoethene	1.784	108	57833	9.306	ppbv	99
15) Acetone	1.839	43	165826	9.033	ppbv	98
16) 1,3-Butadiene	1.609	39	83163	10.093	ppbv	98
17) tert-Butyl alcohol	2.043	59	198776	8.797	ppbv	98
18) 1,1-Dichloroethene	2.036	96	70323	9.192	ppbv	97
19) Isopropyl Alcohol	1.891	45	99004	9.096	ppbv	95
20) Methylene Chloride	2.065	84	62278	8.584	ppbv	99
21) Allyl Chloride	2.098	41	122853	9.664	ppbv	98
22) trans-1,2-Dichloroethene	2.344	96	85089	9.855	ppbv	98
23) Vinyl Acetate	2.467	43	86919	9.552	ppbv #	96
24) 1,1-Dichloroethane	2.412	63	166034	9.942	ppbv	99
25) Ethyl Acetate	2.842	43	536156	10.831	ppbv	100
26) Hexane	2.832	57	210051	10.728	ppbv	95
27) Carbon Disulfide	2.153	76	198444	10.242	ppbv	99
28) Methyl tert-Butyl Ether	2.444	73	110198	7.364	ppbv	99
29) Chloroform	2.855	83	325966	10.532	ppbv	96
30) Cyclohexane	3.865	84	173840	10.236	ppbv	98
31) cis-1,2-Dichloroethene	2.726	61	204611	9.925	ppbv	95
32) 1,1,1-Trichloroethane	3.376	97	305623	10.320	ppbv	99
34) 2-Butanone	2.564	43	302190	11.344	ppbv	99
35) Carbon Tetrachloride	3.768	117	309917	11.778	ppbv	97
36) Benzene	3.664	78	426241	11.094	ppbv	99
37) 1,2-Dichloroethane	3.224	62	235620	10.920	ppbv	99
38) Trichloroethene	4.554	130	167629	10.995	ppbv	97
39) 1,2-Dichloropropane	4.318	63	151232	11.221	ppbv	97
40) 1,4-Dioxane	4.590	88	74021	10.876	ppbv #	100
41) Tetrahydrofuran	3.062	42	169798	11.728	ppbv	98
42) Bromodichloromethane	4.496	83	335595	11.600	ppbv	98
43) Methyl Methacrylate	4.884	69	102724	7.197	ppbv	96
44) 2,2,4-Trimethylpentane	4.648	57	725196	11.670	ppbv	99
45) t-1,3-Dichloropropene	6.422	75	145441	10.886	ppbv	94

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL020325\
 Data File : VL041961.D
 Acq On : 03 Feb 2025 12:57
 Operator : SY/MD
 Sample : VL0203ABS01
 Misc : 400mL/MSVOA_L
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0203ABS01

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 02/04/2025
 Supervised By :Mahesh Dadoda 02/04/2025

Quant Time: Feb 04 02:34:05 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL011425AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug 2

QLast Update : Tue Jan 14 23:15:42 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.609	75	122518	6.985	ppbv	98
47) 1,1,2-Trichloroethane	6.590	97	167038	11.694	ppbv	95
48) Dibromochloromethane	7.390	129	260120	11.803	ppbv	100
49) Bromoform	9.487	173	217089	11.991	ppbv	99
50) 4-Methyl-2-Pentanone	5.765	43	267280	7.350	ppbv	98
51) 2-Hexanone	7.445	43	328817	11.767	ppbv #	96
52) Tetrachloroethene	8.225	164	136143	10.220	ppbv	97
53) Toluene	6.936	91	469033	11.207	ppbv	99
54) 1,2-Dibromoethane	7.655	107	222983	11.356	ppbv	100
56) 1,1,1,2-Tetrachloroethane	8.927	131	214758	11.715	ppbv	99
57) Chlorobenzene	8.924	112	382465	11.284	ppbv	99
58) Ethyl Benzene	9.357	91	646825	11.264	ppbv	98
59) m/p-Xylene	9.555	91	1068374m	23.212	ppbv	
60) o-Xylene	9.966	91	537558	11.766	ppbv	98
61) Styrene	9.872	104	239767	11.699	ppbv	98
62) Isopropylbenzene	10.542	105	802003	11.345	ppbv	99
63) 1,1,2,2-Tetrachloroethane	9.966	83	357905	12.260	ppbv	100
64) n-propylbenzene	10.992	120	208748	11.686	ppbv	99
65) tert-Butylbenzene	11.536	119	731901	11.360	ppbv	99
66) Benzyl Chloride	11.620	91	41013	10.285	ppbv	97
67) sec-Butylbenzene	11.772	105	1015579	11.448	ppbv	99
69) p-Isopropyltoluene	11.931	119	836492	11.451	ppbv	98
70) n-Butylbenzene	12.290	91	824000	11.374	ppbv	100
71) 2-Chlorotoluene	10.905	91	592446	11.329	ppbv	99
72) 4-Ethyltoluene	11.131	105	634050	11.602	ppbv	99
73) 1,3,5-Trimethylbenzene	11.209	105	551677	11.432	ppbv	97
74) 1,2,4-Trimethylbenzene	11.542	105	602744	11.533	ppbv	96
75) 1,3-Dichlorobenzene	11.613	146	348392	11.249	ppbv	99
76) 1,4-Dichlorobenzene	11.678	146	344428	11.367	ppbv	99
77) 1,2-Dichlorobenzene	11.953	146	350456	11.418	ppbv	98
78) Hexachloro-1,3-Butadiene	13.892	225	239328	9.500	ppbv	99
79) Naphthalene	13.523	128	539332	10.249	ppbv	98
80) Naphthalene,2-methyl-	14.468	142	172189	6.902	ppbv	97
81) 1,2,4-Trichlorobenzene	13.452	180	244059	9.502	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

