

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PSEG03

Lab Code: CHEM Case No.: Q1262 SAS No.: Q1262 SDG No.: Q1262

Instrument ID: BNA_F Calibration Date/Time: 02/03/2025 17:14

Lab File ID: BF141373.D Init. Calib. Date(s): 01/29/2025 01/29/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 10:30 17:05

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.303	1.302		-0.1	
Benzaldehyde	0.960	1.038		8.1	
Phenol-d6	1.657	1.598		-3.6	
Phenol	1.757	1.694		-3.6	20.0
bis(2-Chloroethyl)ether	1.347	1.310		-2.7	
2-Chlorophenol	1.396	1.391		-0.4	
2-Methylphenol	1.135	1.109		-2.3	
2,2-oxybis(1-Chloropropane)	2.342	2.192		-6.4	
Acetophenone	0.475	0.486		2.3	
3+4-Methylphenols	1.393	1.420		1.9	
n-Nitroso-di-n-propylamine	0.985	0.981	0.050	-0.4	
Nitrobenzene-d5	0.379	0.382		0.8	
Hexachloroethane	0.571	0.590		3.3	
Nitrobenzene	0.393	0.389		-1.0	
Isophorone	0.656	0.653		-0.5	
2-Nitrophenol	0.185	0.188		1.6	20.0
2,4-Dimethylphenol	0.236	0.237		0.4	
bis(2-Chloroethoxy)methane	0.418	0.408		-2.4	
2,4-Dichlorophenol	0.289	0.293		1.4	20.0
Naphthalene	1.057	1.046		-1.0	
4-Chloroaniline	0.358	0.293		-18.2	
Hexachlorobutadiene	0.194	0.204		5.2	20.0
Caprolactam	0.094	0.096		2.1	
4-Chloro-3-methylphenol	0.310	0.318		2.6	20.0
2-Methylnaphthalene	0.672	0.672		0.0	
Hexachlorocyclopentadiene	0.168	0.186	0.050	10.7	
2,4,6-Trichlorophenol	0.372	0.380		2.2	20.0
2-Fluorobiphenyl	1.299	1.292		-0.5	
2,4,5-Trichlorophenol	0.405	0.413		2.0	
1,1-Biphenyl	1.560	1.536		-1.5	
2-Chloronaphthalene	1.216	1.196		-1.6	
2-Nitroaniline	0.382	0.388		1.6	
Dimethylphthalate	1.351	1.357		0.4	
Acenaphthylene	1.702	1.683		-1.1	
2,6-Dinitrotoluene	0.314	0.315		0.3	
3-Nitroaniline	0.307	0.310		1.0	
Acenaphthene	1.216	1.226		0.8	20.0
2,4-Dinitrophenol	0.146	0.170	0.050	16.4	
4-Nitrophenol	0.225	0.235	0.050	4.4	

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PSEG03

Lab Code: CHEM Case No.: Q1262 SAS No.: Q1262 SDG No.: Q1262

Instrument ID: BNA_F Calibration Date/Time: 02/03/2025 17:14

Lab File ID: BF141373.D Init. Calib. Date(s): 01/29/2025 01/29/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 10:30 17:05

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Dibenzofuran	1.627	1.608		-1.2	
2,4-Dinitrotoluene	0.406	0.423		4.2	
Diethylphthalate	1.310	1.346		2.7	
4-Chlorophenyl-phenylether	0.617	0.621		0.6	
Fluorene	1.262	1.286		1.9	
4-Nitroaniline	0.302	0.327		8.3	
4,6-Dinitro-2-methylphenol	0.118	0.140		18.6	
n-Nitrosodiphenylamine	0.647	0.642		-0.8	20.0
2,4,6-Tribromophenol	0.183	0.194		6.0	
4-Bromophenyl-phenylether	0.224	0.234		4.5	
Hexachlorobenzene	0.238	0.244		2.5	
Atrazine	0.216	0.216		0.0	
Pentachlorophenol	0.139	0.157		12.9	20.0
Phenanthrene	1.075	1.095		1.9	
Anthracene	1.053	1.082		2.8	
Carbazole	0.934	0.977		4.6	
Di-n-butylphthalate	1.132	1.197		5.7	
Fluoranthene	1.062	1.123		5.7	20.0
Pyrene	1.920	1.838		-4.3	
Terphenyl-d14	1.297	1.274		-1.8	
Butylbenzylphthalate	0.671	0.703		4.8	
3,3-Dichlorobenzidine	0.439	0.424		-3.4	
Benzo (a) anthracene	1.380	1.292		-6.4	
Chrysene	1.265	1.240		-2.0	
Bis(2-ethylhexyl)phthalate	0.851	0.932		9.5	
Di-n-octyl phthalate	1.321	1.473		11.5	20.0
Benzo (b) fluoranthene	1.258	1.234		-1.9	
Benzo (k) fluoranthene	1.114	1.155		3.7	
Benzo (a) pyrene	1.063	1.088		2.4	20.0
Indeno (1,2,3-cd) pyrene	1.291	1.353		4.8	
Dibenzo (a,h) anthracene	1.039	1.102		6.1	
Benzo (g,h,i) perylene	1.080	1.119		3.6	
1,2,4,5-Tetrachlorobenzene	0.569	0.571		0.4	
1,4-Dioxane	0.573	0.582		1.6	20.0
2,3,4,6-Tetrachlorophenol	0.319	0.336		5.3	

All other compounds must meet a minimum RRF of 0.010.