

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1262 SAS No.: Q1262 SDG NO.: Q1262

EPA Sample No.: SSTDCCC040 Date Analyzed: 02/03/2025

Lab File ID: BF141373.D Time Analyzed: 17:14

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	104167	6.798	414412	8.08	226326	9.83
UPPER LIMIT	208334	7.298	828824	8.581	452652	10.333
LOWER LIMIT	52083.5	6.298	207206	7.581	113163	9.333
EPA SAMPLE NO.						
01 ETGI-371	81640	6.79	319733	8.08	153664	9.83
02 RBR200030MS	86988	6.79	341693	8.08	181294	9.83
03 RBR200030MSD	76246	6.79	301527	8.08	155603	9.83

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.