

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD011625

Review By	Abdul	Review On	1/17/2025 8:01:10 AM
Supervise By	Ankita	Supervise On	1/17/2025 8:44:47 AM
SubDirectory	PD011625	HP Acquire Method	HP Processing Method pd011625clp

STD. NAME	STD REF.#
Tune/Reschk	PP23732,PP23975
Initial Calibration Stds	PP23731,PP23730,PP23729,PP23728,PP23727,PP23718,PP23724,PP23717,PP23723,PP23716,PP23722,PP23715,PP23721,PP23714,PP23720
CCC	PP23716,PP23722,PP23729
Internal Standard/PEM	
ICV/I.BLK	
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PD087392.D	16 Jan 2025 09:21		AR\AJ	Ok
2	RESC056	RESC011	PD087393.D	16 Jan 2025 09:35		AR\AJ	Ok
3	PEM041	PEM030	PD087394.D	16 Jan 2025 09:49		AR\AJ	Ok
4	TOXAPH101	TOXAPH1042	PD087395.D	16 Jan 2025 10:03		AR\AJ	Ok,M
5	TOXAPH201	TOXAPH2043	PD087396.D	16 Jan 2025 10:17		AR\AJ	Ok
6	TOXAPH301	TOXAPH3044	PD087397.D	16 Jan 2025 10:31		AR\AJ	Ok
7	TOXAPH401	TOXAPH4045	PD087398.D	16 Jan 2025 10:45		AR\AJ	Ok
8	TOXAPH501	TOXAPH5046	PD087399.D	16 Jan 2025 10:59		AR\AJ	Ok
9	INDA101	INDA1047	PD087400.D	16 Jan 2025 11:14		AR\AJ	Ok
10	INDB101	INDB1048	PD087401.D	16 Jan 2025 11:28		AR\AJ	Ok
11	INDA201	INDA2049	PD087402.D	16 Jan 2025 11:42		AR\AJ	Ok
12	INDB201	INDB2050	PD087403.D	16 Jan 2025 11:56		AR\AJ	Ok
13	INDA301	INDA3051	PD087404.D	16 Jan 2025 12:10		AR\AJ	Ok
14	INDB301	INDB3052	PD087405.D	16 Jan 2025 12:24		AR\AJ	Ok
15	INDA401	INDA4053	PD087406.D	16 Jan 2025 12:38		AR\AJ	Ok
16	INDB401	INDB4054	PD087407.D	16 Jan 2025 12:52		AR\AJ	Ok
17	INDA501	INDA5055	PD087408.D	16 Jan 2025 13:06		AR\AJ	Ok
18	INDB501	INDB5056	PD087409.D	16 Jan 2025 13:21		AR\AJ	Ok

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD011625

Review By	Abdul	Review On	1/17/2025 8:01:10 AM
Supervise By	Ankita	Supervise On	1/17/2025 8:44:47 AM
SubDirectory	PD011625	HP Acquire Method	HP Processing Method pd011625clp

STD. NAME	STD REF.#
Tune/Reschk	PP23732,PP23975
Initial Calibration Stds	PP23731,PP23730,PP23729,PP23728,PP23727,PP23718,PP23724,PP23717,PP23723,PP23716,PP23722,PP23715,PP23721,PP23714,PP23720
CCC	PP23716,PP23722,PP23729
Internal Standard/PEM	
ICV/I.BLK	
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	PIBLK092	PIBLK455	PD087410.D	16 Jan 2025 13:35		AR\AJ	Ok
20	PEM042	PEM031	PD087411.D	16 Jan 2025 14:05		AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD020425

Review By	Abdul	Review On	2/5/2025 11:55:04 AM
Supervise By	Ankita	Supervise On	2/5/2025 1:07:36 PM
SubDirectory	PD020425	HP Acquire Method	HP Processing Method pd011625clp

STD. NAME	STD REF.#
Tune/Reschk	PP23732,PP23975
Initial Calibration Stds	PP23731,PP23730,PP23729,PP23728,PP23727,PP23718,PP23724,PP23717,PP23723,PP23716,PP23722,PP23715,PP23721,PP23714,PP23720
CCC	PP23716,PP23722,PP23729
Internal Standard/PEM	
ICV/I.BLK	
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PD087694.D	04 Feb 2025 08:58		AR\AJ	Ok
2	PIBLK116	PIBLK479	PD087695.D	04 Feb 2025 09:12		AR\AJ	Ok,M
3	PEM058	PEM047	PD087696.D	04 Feb 2025 09:26		AR\AJ	Ok,M
4	Q1226-02	DCZJ7	PD087697.D	04 Feb 2025 09:53	need dilution	AR\AJ	Dilution
5	Q1226-02DL	DCZJ7DL	PD087698.D	04 Feb 2025 12:38	need dilution	AR\AJ	Dilution
6	Q1226-02DL2	DCZJ7DL2	PD087699.D	04 Feb 2025 12:52		AR\AJ	Ok,M
7	Q1226-22	DCZL9	PD087700.D	04 Feb 2025 13:15	need dilution	AR\AJ	Dilution
8	Q1226-23	DCZN9	PD087701.D	04 Feb 2025 13:29	TCMX high in 2nd column	AR\AJ	Ok,M
9	Q1226-24	DCZN2	PD087702.D	04 Feb 2025 13:43		AR\AJ	Ok,M
10	Q1226-25	DCZN7	PD087703.D	04 Feb 2025 13:57	DCB low in 1st column	AR\AJ	Ok,M
11	Q1226-26	DCZP4	PD087704.D	04 Feb 2025 14:11		AR\AJ	Ok,M
12	Q1268-11	PEST-GPC2-BLANK	PD087705.D	04 Feb 2025 14:26		AR\AJ	Ok
13	Q1268-12	PEST-GPC2-BLANK-S	PD087706.D	04 Feb 2025 14:40		AR\AJ	Ok,M
14	PB166476BL	PBLK476	PD087707.D	04 Feb 2025 15:04		AR\AJ	Ok,M
15	Q1202-11	E29A2	PD087708.D	04 Feb 2025 15:18		AR\AJ	Ok
16	Q1226-08DL	DCZP0DL	PD087709.D	04 Feb 2025 15:32		AR\AJ	Ok,M
17	Q1226-09DL	DCZP1DL	PD087710.D	04 Feb 2025 15:46	need further dilution	AR\AJ	Dilution
18	PIBLK117	PIBLK480	PD087711.D	04 Feb 2025 16:07		AR\AJ	Ok,M

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD020425

Review By	Abdul	Review On	2/5/2025 11:55:04 AM
Supervise By	Ankita	Supervise On	2/5/2025 1:07:36 PM
SubDirectory	PD020425	HP Acquire Method	HP Processing Method pd011625clp
STD. NAME	STD REF.#		
Tune/Reschk	PP23732,PP23975		
Initial Calibration Stds	PP23731,PP23730,PP23729,PP23728,PP23727,PP23718,PP23724,PP23717,PP23723,PP23716,PP23722,PP23715,PP23721,PP23714,PP23720		
CCC	PP23716,PP23722,PP23729		
Internal Standard/PEM			
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	INDA309	INDA3221	PD087712.D	04 Feb 2025 16:21		AR\AJ	Ok,M
20	INDB309	INDB3222	PD087713.D	04 Feb 2025 16:35		AR\AJ	Ok,M
21	Q1202-05	E2979	PD087714.D	04 Feb 2025 16:49		AR\AJ	Ok
22	Q1202-07	E2977	PD087715.D	04 Feb 2025 17:03		AR\AJ	Ok
23	Q1202-10	E29A1	PD087716.D	04 Feb 2025 17:30		AR\AJ	Ok
24	Q1202-14	E29A5	PD087717.D	04 Feb 2025 17:44		AR\AJ	Ok
25	Q1202-04	E2978	PD087718.D	04 Feb 2025 17:58		AR\AJ	Ok
26	PB166544BL	PBLK544	PD087719.D	04 Feb 2025 18:12		AR\AJ	Ok
27	PB166544BS	PLCS544	PD087720.D	04 Feb 2025 18:27		AR\AJ	Ok,M
28	Q1229-02	E2991	PD087721.D	04 Feb 2025 18:41		AR\AJ	Ok
29	Q1229-04	E2993	PD087722.D	04 Feb 2025 18:55		AR\AJ	Ok
30	Q1229-05MS	E2993MS	PD087723.D	04 Feb 2025 19:09		AR\AJ	Ok,M
31	Q1229-06MSD	E2993MSD	PD087724.D	04 Feb 2025 19:23		AR\AJ	Ok,M
32	Q1229-09	E29A9	PD087725.D	04 Feb 2025 19:37		AR\AJ	Ok
33	Q1229-11	E29B5	PD087726.D	04 Feb 2025 19:51	DCB Low in both column, FB	AR\AJ	Ok,NS
34	PIBLK118	PIBLK481	PD087727.D	04 Feb 2025 20:05		AR\AJ	Ok
35	PEM059	PEM048	PD087728.D	04 Feb 2025 20:19		AR\AJ	Ok,M
36	Q1229-12	E29B8	PD087729.D	04 Feb 2025 20:47		AR\AJ	Ok
37	Q1229-13	E29B9	PD087730.D	04 Feb 2025 21:01		AR\AJ	Ok

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD020425

Review By	Abdul	Review On	2/5/2025 11:55:04 AM
Supervise By	Ankita	Supervise On	2/5/2025 1:07:36 PM
SubDirectory	PD020425	HP Acquire Method	HP Processing Method pd011625clp
STD. NAME	STD REF.#		
Tune/Reschk	PP23732,PP23975		
Initial Calibration Stds	PP23731,PP23730,PP23729,PP23728,PP23727,PP23718,PP23724,PP23717,PP23723,PP23716,PP23722,PP23715,PP23721,PP23714,PP23720		
CCC	PP23716,PP23722,PP23729		
Internal Standard/PEM			
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

38	Q1229-14	E29C0	PD087731.D	04 Feb 2025 21:15		AR\AJ	Ok
39	Q1229-16	E29C2	PD087732.D	04 Feb 2025 21:29		AR\AJ	Ok,M
40	Q1229-17	E29C3	PD087733.D	04 Feb 2025 21:44		AR\AJ	Ok
41	Q1229-18	E29C4	PD087734.D	04 Feb 2025 21:58		AR\AJ	Ok
42	PIBLK119	PIBLK482	PD087735.D	04 Feb 2025 22:12		AR\AJ	Ok
43	INDA310	INDA3223	PD087736.D	04 Feb 2025 22:26		AR\AJ	Ok
44	INDB310	INDB3224	PD087737.D	04 Feb 2025 22:40		AR\AJ	Ok,M

M : Manual Integration