

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1268 SAS No.: Q1268 SDG NO.: Q1268

EPA Sample No.: SSTD020796 Date Analyzed: 02/05/2025

Lab File ID: BP023937.D Time Analyzed: 15:43

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	423241	7.769	1842950	10.55	1130160	14.39
UPPER LIMIT	846482	8.269	3685900	11.046	2260320	14.887
LOWER LIMIT	211621	7.269	921475	10.046	565080	13.887
EPA SAMPLE NO.						
01 SVOC-GPC2-BLANK	360926	7.77	1569510	10.54	997207	14.39

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.