

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX011525

Review By	John Carlone	Review On	1/15/2025 4:32:35 PM
Supervise By	Maresh Dadoda	Supervise On	1/16/2025 7:35:39 AM
SubDirectory	VX011525	HP Acquire Method	HP Processing Method 82X011525W.M

STD. NAME	STD REF.#
Tune/Reschk	VP132553
Initial Calibration Stds	VP132558,VP132559,VP132560,VP132561,VP132562,VP132563
CCC	
Internal Standard/PEM	
ICV/I.BLK	VP132564
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX044647.D	15 Jan 2025 09:30		JC/MD	Ok
2	VSTDIC001	VSTDIC001	VX044648.D	15 Jan 2025 10:31		JC/MD	Ok,M
3	VSTDIC005	VSTDIC005	VX044649.D	15 Jan 2025 11:17		JC/MD	Ok,M
4	VSTDIC020	VSTDIC020	VX044650.D	15 Jan 2025 11:40		JC/MD	Ok,M
5	VSTDIC050	VSTDIC050	VX044651.D	15 Jan 2025 12:02	Method failed for Acrolein	JC/MD	Ok,M
6	VSTDIC100	VSTDIC100	VX044652.D	15 Jan 2025 13:14		JC/MD	Ok,M
7	VSTDIC150	VSTDIC150	VX044653.D	15 Jan 2025 13:37		JC/MD	Ok,M
8	IBLK	IBLK	VX044654.D	15 Jan 2025 14:00		JC/MD	Ok
9	VSTDICV050	ICVVX011525	VX044655.D	15 Jan 2025 15:15		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX020525

Review By	Mahesh Dadoda	Review On	2/6/2025 1:23:36 PM
Supervise By	Semsettin Yesilyurt	Supervise On	2/6/2025 1:25:11 PM
SubDirectory	VX020525	HP Acquire Method	HP Processing Method 82X011525W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP132888
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP132889,VP132890

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX044832.D	05 Feb 2025 08:02		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX044833.D	05 Feb 2025 09:51	V13516	JC/MD	Ok,M
3	VX0205MBL01	VX0205MBL01	VX044834.D	05 Feb 2025 10:18		JC/MD	Ok
4	VX0205WBL01	VX0205WBL01	VX044835.D	05 Feb 2025 10:41		JC/MD	Ok
5	VX0205WBS01	VX0205WBS01	VX044836.D	05 Feb 2025 11:59	Surrogate Fail	JC/MD	Not Ok
6	IBLK	IBLK	VX044837.D	05 Feb 2025 12:22		JC/MD	Ok
7	VX0205WBS02	VX0205WBS02	VX044838.D	05 Feb 2025 13:08		JC/MD	Ok,M
8	Q1268-03	Storage-Blank-WATER	VX044839.D	05 Feb 2025 13:31	vial B pH<2	JC/MD	Ok
9	Q1268-04	Storage-Blank-SAMPLE	VX044840.D	05 Feb 2025 13:54	vial B pH<2	JC/MD	Ok
10	Q1250-01	BP-VPB-192-TB-20250	VX044841.D	05 Feb 2025 14:17	END CCC fail for comp.#06; TB	JC/MD	ReRun
11	Q1268-01	Storage-Blank-SOIL-RE	VX044842.D	05 Feb 2025 14:40	vial B pH<2	JC/MD	Ok
12	Q1268-02	Storage-Blank-WATER	VX044843.D	05 Feb 2025 15:04	vial B pH<2	JC/MD	Ok
13	Q1250-04	BP-VPB-192-GW-320-3	VX044844.D	05 Feb 2025 15:27	END CCC fail for comp.#06	JC/MD	ReRun
14	Q1250-05	BP-VPB-192-GW-340-3	VX044845.D	05 Feb 2025 15:50	END CCC fail for comp.#06	JC/MD	ReRun
15	Q1250-10	BP-VPB-192-GW-380-3	VX044846.D	05 Feb 2025 16:13	END CCC fail for comp.#06	JC/MD	ReRun
16	Q1250-11	BP-VPB-192-GW-400-4	VX044847.D	05 Feb 2025 16:36	END CCC fail for comp.#06	JC/MD	ReRun
17	Q1250-12	BP-VPB-192-GW-440-4	VX044848.D	05 Feb 2025 16:59	END CCC fail for comp.#06	JC/MD	ReRun
18	Q1250-02	BP-VPB-192-GW-420-4	VX044849.D	05 Feb 2025 17:22	END CCC fail for comp.#06	JC/MD	ReRun

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX020525

Review By	Mahesh Dadoda	Review On	2/6/2025 1:23:36 PM
Supervise By	Semsettin Yesilyurt	Supervise On	2/6/2025 1:25:11 PM
SubDirectory	VX020525	HP Acquire Method	HP Processing Method 82X011525W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP132888		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP132889,VP132890		

19	Q1250-03	BP-VPB-192-GW-300-3	VX044850.D	05 Feb 2025 17:45	END CCC fail for comp.#06	JC/MD	ReRun
20	Q1250-06	BP-VPB-192-GW-360-3	VX044851.D	05 Feb 2025 18:09	END CCC fail for comp.#06	JC/MD	ReRun
21	Q1250-07MS	BP-VPB-192-GW-360-3	VX044852.D	05 Feb 2025 18:32	END CCC fail for comp.#06	JC/MD	ReRun
22	Q1250-08MSD	BP-VPB-192-GW-360-3	VX044853.D	05 Feb 2025 18:55	END CCC fail for comp.#06	JC/MD	ReRun
23	VSTDCCC050	VSTDCCC050EC	VX044854.D	05 Feb 2025 19:17	Failed for comp.#06	JC/MD	Ok,M

M : Manual Integration