

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CHEM02
 Lab Code: CHEM Case No.: Q1268 SAS No.: Q1268 SDG NO.: Q1268
 Lab File ID: VX044833.D Date Analyzed: 02/05/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:51
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	92963	12.018				
UPPER LIMIT	185926	12.518				
LOWER LIMIT	46481.5	11.518				
EPA SAMPLE NO.						
Storage-Blank-SOIL-REF-6	67577	12.02				
Storage-Blank-WATER-REF-5	68345	12.02				
Storage-Blank-WATER-REF-4	76428	12.02				
Storage-Blank-SAMPLE-MANAGEMENT	77528	12.02				
VX0205WBL01	90416	12.02				
VX0205WBS02	86992	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.