

---

**From:** DeBerry, Eric <Eric.Deberry@gdit.com>  
**Sent:** Monday, February 03, 2025 9:30 AM  
**To:** Deepak Parmar; Sohil Jodhani; Mohammad Ahmed  
**Cc:** Bauer, Heather E; Johnson, Matthew; Burman, Jarmael; 'Moody, Brett'; Gambrah, Derrick; Patel, Bhavita; Vargas.Magda@epa.gov; Britz, Helen  
**Subject:** Task Area SST | Region 03 | Case 51988 | Lab ACE | Issue Insufficient/inappropriate designation of laboratory QC | FINAL  
**Attachments:** SKM\_95825020111380.pdf

EXTERNAL EMAIL - This email was sent by a person from outside your organization. Exercise caution when clicking links, opening attachments or taking further action, before validating its authenticity.

Secured by Check Point

Good morning Deepak,

Issue: Laboratory QC is scheduled for ICP-AES and TCLP ICP-AES analysis, but no sample was designated for laboratory QC on the COC. The laboratory would like to use sample MBHHA2 for laboratory QC. The laboratory confirms that these samples are not blanks, rinsates, or PT.

Resolution: Per SOW SFAM01.1 Exhibit A, Section 5.5.4.1., please note the issue in the SDG Narrative and proceed with the analysis of the samples.

Please note that the laboratory may contact the appropriate CLP PM should any defects need to be waived for this issue.

Thanks,

Eric DeBerry

Associate Environmental Analyst  
CLP QSS Coordinator – EPA Regions 1 & 3

*Under contract to the EPA*

T: (571) 833-5166  
[Eric.DeBerry@GDIT.com](mailto:Eric.DeBerry@GDIT.com)  
15036 Conference Center Drive  
Chantilly, VA 20151  
[www.gdit.com](http://www.gdit.com)

**GENERAL DYNAMICS**  
Information Technology Corp.

---

**From:** Deepak Parmar <[Deepak.Parmar@alliancetg.com](mailto:Deepak.Parmar@alliancetg.com)>  
**Sent:** Monday, February 3, 2025 8:21 AM  
**To:** DeBerry, Eric <[Eric.Deberry@gdit.com](mailto:Eric.Deberry@gdit.com)>

Cc: Sohil Jodhani <[Sohil.Jodhani@AllianceTG.com](mailto:Sohil.Jodhani@AllianceTG.com)>

Subject: Region 3 | Case 51988 | Lab ACE | Issue Discrepancies with tags, jars, and/or COC/QC

**This Message Is From an External Sender**

Please use caution with links, attachments, and any requests for credentials.

Good morning,

Issue 1: QC Scheduled for ICP-AES and TCLP ICP-AES analysis However, a sample was not designated for Laboratory QC. Lab like to use sample MBHHA2 for Lab QC. these samples is not blanks, rinsates or PT.

Please see attachment for your reference.

Thanks & Regards,



**Deepak Parmar**

QA/QC

**An Alliance Technical Group Company**

Main: 908-789-8900

Direct: 908-728-3154

Address: 284 Sheffield St, Ste 1, Mountainside, NJ 07092

[www.alliancetg.com](http://www.alliancetg.com)

