

FORM 10B-IN
ICP-AES INTERELEMENT CORRECTION FACTORS

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51988 MA No.: SDG No.: MBHHA1
Instrument ID: P5 Date : 03/15/2024

Analyte	Wave- Length (nm)	Interelement Correction Factors For				
		Cr	Cu	Mn	Mo	Ni
Arsenic	189.042	0	0	0	0	0
Barium	493.409	0	0	0	0	0
Cadmium	214.438	0	0	0	0	0
Chromium	267.716	0	0	0.00016	0	0
Lead	220.353	0	0.0014	0.00034	-0.00148	0.00063
Selenium	196.09	0	0	0.00047	0	0
Silver	328.068	0	0	0.000146	0	0

FORM 10B-IN
ICP-AES INTERELEMENT CORRECTION FACTORS

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51988 MA No.: SDG No.: MBHHA1
Instrument ID: P5 Date : 03/15/2024

Analyte	Wave- Length (nm)	Interelement Correction Factors For				
		Pb	Ti	V	Zn	
Arsenic	189.042	0	0	0	0	
Barium	493.409	0	0	0	0	
Cadmium	214.438	0	0	0	0	
Chromium	267.716	0	0	0	0	
Lead	220.353	0	0	0	0	
Selenium	196.09	0	0	0	0	
Silver	328.068	0	0	-0.000889	0	

FORM 10B-IN
ICP-AES INTERELEMENT CORRECTION FACTORS

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51988 MA No.: SDG No.: MBHHA1
Instrument ID: P5 Date : 03/15/2024

Analyte	Wave- Length (nm)	Interelement Correction Factors For				
Arsenic	189.042	0	0	0	0	
Barium	493.409	0	0	0	0	
Cadmium	214.438	0	0	0	0	
Chromium	267.716	0	0	0	0	
Lead	220.353	0	0	0	0	
Selenium	196.09	0	0	0	0	
Silver	328.068	0	0	0	0	