

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51988 MA No. : SDG No.: MBHHA1Initial Calibration Verification Source : MP84389Continuing Calibration Verification Source : MP84392Run Batch: LB134617 Analytical Method: ICP-AESConcentration Units: $\mu\text{g/L}$

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV002				ID: CCV003				ID: CCV004		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Arsenic	1000	1100	108	0	5000	4900	98	0	4900	99	0
Barium	520	520	101	3	10000	10000	103	1	10000	102	0
Cadmium	510	490	96	0	2500	2600	102	0	2500	102	0
Chromium	520	550	106	0	15500	16000	102	0	16000	101	0
Lead	1000	990	99	0	25000	25000	101	0	25000	101	0
Selenium	1000	1000	100	0	5000	5000	99	0	5000	100	0
Silver	250	250	98	1	1250	1300	101	0	1300	100	0

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51988 MA No. : SDG No.: MBHHA1Initial Calibration Verification Source : MP84389Continuing Calibration Verification Source : MP84392Run Batch: LB134644 Analytical Method: ICP-AESConcentration Units: $\mu\text{g/L}$

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV003				ID: CCV005				ID: CCV006		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Arsenic	1000	1100	107	0	5000	5000	100	0	5000	99	1
Barium	520	520	101	1	10000	10000	101	0	10000	103	0
Cadmium	510	480	94	0	2500	2500	101	0	2600	102	0
Chromium	520	550	105	0	15500	15000	100	0	16000	101	0
Lead	1000	980	98	0	25000	25000	100	0	25000	101	0
Selenium	1000	1000	100	0	5000	4900	99	0	5000	100	0
Silver	250	240	97	0	1250	1200	100	0	1300	102	0