

FORM 4 - IN
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51988 MA No. : SDG No.: MBHHA1
 Analytical Method: ICP-AES ICSA Source: EPA-1211
 Instrument ID: P5 ICSB Source: EPA-0710
 Run Batch: LB134617
 Concentration Units: ug/L

Analyte	True		Found			
	ICSA	ICSAB	ICSA	%R	ICSAB	%R
Aluminum	255000	247000	240000	94	240000	95
Antimony	0	618	3.9		540	87
Arsenic	0	104	-2.4		97	93
Barium	6	537	7.0	117	520	97
Beryllium	0	495	2.6		510	103
Cadmium	1	972	0.93	93	980	101
Calcium	245000	235000	250000	102	250000	105
Chromium	52	542	52	101	590	108
Cobalt	0	476	30		520	109
Copper	2	511	-14	-695	490	96
Iron	101000	99300	100000	100	99000	100
Lead	0	49	1.8		51	103
Magnesium	255000	248000	250000	97	240000	98
Manganese	7	507	6.1	87	510	100
Nickel	2	954	7.1	352	980	103
Potassium	0	0	-1000		-2500	
Selenium	0	46	-26		25	55
Silver	0	201	-8.3		200	100
Sodium	0	0	110		250	
Thallium	0	108	-3.2		97	90
Vanadium	0	491	-0.22		490	101
Zinc	0	952	-34		900	94

FORM 4 - IN
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51988 MA No. : SDG No.: MBHHA1
 Analytical Method: ICP-AES ICSA Source: EPA-1211
 Instrument ID: P5 ICSB Source: EPA-0710
 Run Batch: LB134644
 Concentration Units: ug/L

Analyte	True		Found			
	ICSA	ICSAB	ICSA	%R	ICSAB	%R
Aluminum	255000	247000	240000	92	240000	96
Antimony	0	618	16		570	92
Arsenic	0	104	-0.26		99	95
Barium	6	537	2.8	47	520	96
Beryllium	0	495	3.0		520	104
Cadmium	1	972	2.2	223	990	102
Calcium	245000	235000	240000	99	250000	104
Chromium	52	542	53	102	580	108
Cobalt	0	476	29		520	109
Copper	2	511	-14	-707	490	96
Iron	101000	99300	99000	98	100000	101
Lead	0	49	-0.31		46	94
Magnesium	255000	248000	240000	95	250000	99
Manganese	7	507	6.2	88	520	103
Nickel	2	954	4.6	228	990	104
Potassium	0	0	-200		-510	
Selenium	0	46	-28		19	42
Silver	0	201	-4.9		200	100
Sodium	0	0	-100		-250	
Thallium	0	108	-1.7		97	90
Vanadium	0	491	3.5		510	104
Zinc	0	952	-31		920	97