

Lab Name:	Chemtech Consulting Group	Contract:	EPW14030
Lab Code:	ACE	Case No.:	51988
MA No. :		SDG No.:	MBHHA1
Matrix :	Water	Analytical Method:	ICP-AES
% Solids:			

Concentration Units (µg/L, mg/L or mg/kg dry weight): ug/L

Analyte	Control Limit %R	Spiked Sample Result (SSR) Q	Sample Result (SR) Q	Spike Added (SA)	%R	Q
Arsenic	75 - 125	46	10.0 U	40.0	115	
Barium	75 - 125	3900	1900	2000	100	
Cadmium	75 - 125	55	2.8 J	50.0	105	
Chromium	75 - 125	210	4.0 J	200	105	
Lead	75 - 125	46	23	20.0	114	
Selenium	75 - 125	95	35.0 U	100	95	
Silver	75 - 125	38	10.0 U	50.0	75	