

Initial and Continuing Calibration Verification

Client: METEM A GE POWER Business

SDG No.: Q1278

Project: Parsippany Wastewater Quarterly 2025

RunNo.: LB134576

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Cyanide	ICV1	mg/L	0.097	0.099	98	85-115	02/05/2025
Sample ID: Cyanide	CCV1	mg/L	0.25	0.25	100	90-110	02/05/2025
Sample ID: Cyanide	CCV2	mg/L	0.25	0.25	100	90-110	02/05/2025

Initial and Continuing Calibration Verification

Client: METEM A GE POWER Business

SDG No.: Q1278

Project: Parsippany Wastewater Quarterly 2025

RunNo.: LB134639

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
Field pH		pH	7.00	7	100	90-110	02/03/2025
Sample ID:	CCV1						
Field pH		pH	7.00	7.00	100	90-110	02/03/2025
Sample ID:	CCV2						
Field pH		pH	7.00	7.00	100	90-110	02/03/2025