

Laboratory Control Sample Summary

Client: METEM A GE POWER Business

SDG No.: Q1278

Project: Parsippany Wastewater Quarterly 2025

Run No.: LB134576

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB166577BS							
Cyanide	mg/L	0.1	0.096		96	1	85-115	02/05/2025